

Performance comparison of two malaria rapid diagnostic test with real time polymerase chain reaction and gold standard of microscopy detection method

by Puspa Wardhani

Submission date: 19-Jul-2021 11:28AM (UTC+0800)

Submission ID: 1621362710

File name: in_reaction_and_gold_standard_of_microscopy_detection_method.pdf (1.61M)

Word count: 4517

Character count: 24084

Performance comparison of two malaria rapid diagnostic test with real time polymerase chain reaction and gold standard of microscopy detection method

Puspa Wardhani,^{1,5} Trieva Verawaty Butarbutar,^{2,5} Christophorus Oetama Adiatmaja,^{2,5} Amarensi Milka Betaubun,^{3,5} Nur Hamidah,⁴ Aryati^{1,5}

¹Clinical Pathology Department; ²Clinical Pathology Specialist Study Programme, Clinical Pathology Department; ³Clinical Pathology Subspecialist Study Programme, Clinical Pathology Department, Faculty of Medicine; ⁴Science and Technology Faculty, C Campus, Universitas Airlangga; ⁵Dr. Soetomo General Academic Hospital, Surabaya, Indonesia

Abstract

Background: The diagnostic test for malaria is mostly based on Rapid Diagnostic Test (RDT) and detection by microscopy. Polymerase Chain Reaction (PCR) is also a sensitive detection method that can be considered as a diagnostic tool. The outcome of malaria microscopy detection depends on the examiner's ability and experience. Some RDT has been distributed in Indonesia, which needs to be evaluated for their results.

Objective: This study aimed to compare the performance of RightSign RDT and ScreenPlus RDT for detection of *Plasmodium* in human blood. We used specific real-time polymerase chain reaction (abTESTMalaria qPCRII) and gold standard of microscopy detection method to measure diagnostic efficiency.

Methods: Blood specimens were evaluated using RightSign RDT, ScreenPlus RDT, Microscopy detection, and RT-PCR as the protocol described. The differences on specificity (Sp), sensitivity (Sn), positive predictive value (PPV), and negative predictive value (NPV) were analyzed using McNemar and Kruskal Wallis analysis.

Results: A total of 105 subjects were recruited. Based on microscopy test, RightSign RDT had sensitivity, Specificity, PPV, NPV, 100%, 98%, 98.2%, 100%, respectively. ScreenPlus showed 100% sensitivity, 98% specificity, 98.2% PPV, 100% NPV. The sensitivity of both RDTs became lower (75%) and the specificity

higher (100 %) when using real-time PCR. Both RDTs showed a 100% agreement. RT-PCR detected higher mix infection when compared to microscopy and RDTs.

Conclusion: RightSign and ScreenPlus RDT have excellent performance when using microscopy detection as a gold standard. Real-time PCR method can be considered as a confirmation tool for malaria diagnosis.

Introduction

Indonesia is a malaria endemic area. The province of Papua has Indonesia's highest malaria burden. All *Plasmodium* species are present in Papua, including the *Plasmodium knowlesi* (*P. knowlesi*) originally discovered on Kalimantan Island. The most common types of Plasmodium infection in Papua are *Plasmodium falciparum* (*P. falciparum*) and *Plasmodium vivax* (*P. vivax*). *Plasmodium ovale* (*P. ovale*) and *Plasmodium malariae* (*P. malariae*) also can be found in Papua. The highest cause of morbidity and mortality in Papua is *P. falciparum*.^{1,2}

Several methods for diagnosing malaria have been established since WHO has confirmed how important new tests are to diagnose *Plasmodium* spp rapidly, reliably, accurately and cheaply, to address numerous shortcomings in microscopic examination as the WHO's gold standard for malaria research.^{3,4}

Microscopic examination has advantages, namely allowing to identify species definitively, determine the condition of parasitemia, monitor malaria treatment response, easy execution and inexpensive.^{3,5} However, the weakness of microscopy examination is the difficulties in detecting extremely low parasitemia and mix infections, and it is time consuming. Although microscopy examination is gold standard in malaria diagnosis, it is better to be accompanied by RDTs or other methods.^{3,5}

Other testing methods have been developed to diagnose malaria are the based on proteins produced by *Plasmodium*, such as Histidine-Rich Protein 2 (HRP-2) or enzymes such as Pan-Plasmodium Lactate Dehydrogenase (pLDH) and Pan-Specific Aldolase. This method has been widely used and is considered an alternative to malaria testing. This examination is also very prominent, especially in areas where microscopic examination is not available, or it is available but there is no laboratory staff experienced in the examination of blood or some other disadvantage of microscopy examination. Commercially enzyme-based

Correspondence: Puspa Wardhani, Department of Clinical Pathology, Faculty of Medicine, Universitas Airlangga/Dr. Soetomo. General Academic Hospital, Jl. Prof. Dr. Moestopo No. 47, 60132, Surabaya, Indonesia.
Tel: +628123176937
E-mail: puspa_pk@yahoo.co.id,

Key words: Malaria, rapid diagnostic test, real-time polymerase chain reaction, microscopy detection.

Contributions: Conceptualization: PW. Data curation: PWTVB COAAMB. Formal analysis: PW AA NH. Funding: KEMENRIS-TEKDIKTI RI. Investigation: PW PW TVB COA AMB, NH, AA. Methodology: PW AA. Project administration: PW TVB AMB COA Resources: PW AA NH. Supervision: PW AA NH. Validation: PW AA. Writing-original draft, review & editing: PW.

Conflicts of interest: The authors declare no potential conflict of interest.

Funding: The work was supported by Simlitabmas Programme by Kementerian Riset Teknologi dan Pendidikan Tinggi Republik Indonesia (KEMENRISTEKDIKTI RI), contract number: 544/UN3.14/LT/2019.

Acknowledgements: The authors acknowledge the funding research support of Simlitabmas programme and Merauke Hospital, Papua for sampling place support, Parasitology department Faculty of Medicine, Brawijaya University for consultation in laboratory method.

Conference presentation: The paper had been presented at INSBIOMM International conference on latest perspectives on Infectious Diseases, including Biotreats and Military Medicine, in August 27-28th, 2019, Surabaya, INDONESIA.

Received for publication: 17 February 2020.
Accepted for publication: 1 July 2020.

This work is licensed under a Creative Commons Attribution-NonCommercial 4.0 International License (CC BY-NC 4.0).

©Copyright: the Author(s), 2020
Licensee PAGEPress, Italy
Infectious Disease Reports 2020; 12(s1):8731
doi:10.4081/idr.2020.8731

examination tools generally use serological techniques to detect antigens, such as Rapid Diagnostic Tests (RDTs) or Immunochromatography Test (ICT) and Enzyme Linked Immunosorbent Assay (ELISA).^{6,7}

The RDTs or ICT detect malaria

antigens when the blood of the patient passes through a membrane containing Plasmodium-specific antibodies. The benefits of RDT are reliability in testing, quick processing times and tests, and cheap review fees, which is why it is very useful in endemic areas. The most commonly used RDTs are methods of detection based on HRP2 and pLDH. The advantages of RDTs are their parasitic detection limit of 50-100 parasites/ μL . The disadvantage of RDT are reduced effectiveness and efficiency in examining large samples, because time consuming and a high level of concentration and accuracy are required.^{3,8}

Most RDTs examine two proteins at once, such as a combination of a specific protein *P. falciparum* (HRP-2) with pLDH, a protein that is shared by all species of Plasmodium (non-specific) pLDH) or HRP-2 combination combined with a species specific protein species of *Plasmodium vivax* (Pv-pLDH). Three proteins also can be detected at once, such as HRP-2, non-specific pLDH and Pv-pLDH that can distinguish between *falciparum* malaria and non-*falciparum* malaria or mixed malaria.⁹

RightSign and ScreenPlus are available commercially in Indonesia. RightSign detects *P. falciparum* (HRP-2) and Pv-pLDH. ScreenPlus detects *P. falciparum* (HRP-2) and Pan LDH. Both RDTs have not been evaluated in field study in Indonesia.

Molecular-based Polymerase Chain Reaction (PCR) techniques have used for years in detecting *Plasmodium spp.* A study on Polymerase Chain Reaction (PCR) even said that the sensitivity and specificity of PCR are higher than microscopic examination. PCR has a high sensitivity and specificity, but has several disadvantages: it is a fairly complex examination method, it is expensive, and requires expertise to be performed.^{3,10,11}

This study aimed to compare the performance of RightSign RDT and ScreenPlus RDT for detection *Plasmodium* in human blood. The difference performance of RDTs malaria compared to microscopy detection, and Real-Time Polymerase Chain Reaction (Real Time (RT)-PCR abTES™Malaria qPCR II).

Materials and Methods

Ethical approval

Ethical approval number 169/EC/KEPK/FKUA/2019 was obtained from the Health Research Ethics Committee of the Faculty of Medicine, Universitas Airlangga.

Sample collection

This study was a cross-sectional analytic study conducted at Merauke Hospital, Papua and Clinical Pathology Laboratory, Faculty of Medicine, Universitas Airlangga /Dr. Soetomo General Academic Hospital Surabaya during November 2018-June 2019. A total of 105 whole blood samples were collected in tubes containing Ethylene Diamine Tetraacetic Acid (EDTA).

Species identification and determination of parasite density

Identification of Plasmodium species and calculation of the parasitemia index (PI) were performed microscopically on Giemsa-stained thick and thin blood smears. PI was calculated using WHO guidelines Simultaneously the samples were using RightSign RDT and ScreenPlus RDT.¹²

RightSign RDT and ScreenPlus RDT

The positivity and antigen detection of the Plasmodium species in RightSign RDT were obtained through lines or bands that arise in the line test. RightSign uses the immunochromatography test (ICT) method, with nitrocellulose membranes detects *P. falciparum*-specific anti-Histidine Rich Protein II (HRP II) on the Pf test line and non-specific anti-Plasmodium Lactate Dehydrogenase (pLDH) from *Plasmodium spp* (*P. falciparum*, *P. vivax*, *P. malariae*, *P. ovale*) on the Pan test line. ScreenPlus detected *P. vivax* specific anti-Plasmodium Lactate Dehydrogenase (pLDH) antibody on the Pv test line and *P. falciparum*-specific anti-Histidine Rich Protein II (HRP II) antibodies. The differences on specificity (Sp), sensitivity (Sn), positive predictive value (PPV), and negative predictive value (NPV) were analyzed using McNemar and Kruskal Wallis analysis.

Results confirmation

Results confirmation of thick and thin blood smears examination was carried out in the Clinical Pathology Laboratory Faculty of Medicine, Universitas Airlangga and Dr. Soetomo general academic hospital, Surabaya. The inclusion criteria for sample acceptance were patients of all ages, males and females, clinical examinations showed symptoms of fever (specific or non-specific malaria) and who were willing to take part in the study by signing an informed consent form. The exclusion criteria were patients who have received malaria treatment and patients with fever but with negative results of malaria microscopy examination. The interpretation of microscopy and RDTs

results was carried out by two blind and independent observers. Based on the McNemar test between the two observers' results, the results were not significantly different with $P < 0.05$.

Real Time PCR

DNA examination of Plasmodium malaria using the Rotor-Gene® Q PCR from Qiagen, Tokyo, Japan with the abTESTMMalaria qPCR III reagent kit. AbTESTMMalaria qPCR III (AITbiotech Pte Ltd, Singapore) can identify four species of Plasmodium spp. (*P. falciparum*, *P. vivax*, *P. ovale*, *P. malariae*, *P. knowlesi*). There are two main steps: DNA extraction from blood samples and amplification of DNA extracts using primers pairs and probes that hydrolyze double-dye (double-dye hydrolysis probes) which are very specific. Double-dye hydrolysis probes are fluorescent substances that are able to emit light, then the light will be captured by optical detectors on the Rotor-Gene® Q device, which results in an increase in signal in the wave graph. This probe attaches to the primers of each Plasmodium species, which are compatible with the Rotor-Gene® Q optical detector channel, each consisting of FAM (compatible with Green channels), HEX (compatible with Yellow channels), ROX (compatible with Orange channels) and Cy5 (compatible with Red channel), Quasar 705 (compatible with Crimson channel), VIC, TAMRA, TEXAS RED.

Results

Acquisition of samples

As many as 105 samples were obtained, which consisted of 67/105 (63.8%) were male and 38/105 (36.2%) were female. The average of age of patients was 31.19 years old. Median and long IQR fever obtained from 28 malaria patients who had a history of fever were five days (2-60) Microscopy examination resulted in 54 (51.42%) out of 105 samples were positive of Plasmodium, and 51 (48.58%) were negative. Examination using RightSign and ScreenPlus resulted in 51.42% were positive (Table 1).

Parasitemia index

Out of 54 positive samples, 34 (63%) were *P. falciparum*, 17(31.5%) were *P. vivax* and 3 (5.5%) were mix infections of both species. Based on parasite density, the highest parasitemia index was *P. vivax* which in range of 1000-10,000/ μL , followed by *P. falciparum* of 1000-10,000/ μL .

and 10,001-200,000/ μ L, while the mixed infection was evenly distributed as <1000, 1000- 10,000 and 10,001-200,000/ μ L.

In this study, a high parasitemia index was obtained with each dominance of 1000-10,000 parasites/ μ L was 41.2% in *P.falciparum* and 52.9% in *P.vivax*, while the parasitemia index <1000 parasites/ μ L was 5.9% in *P.falciparum* and *P.vivax* (Table 1).

The negative samples

Fifty negative Plasmodium based on RDT examination, consisted of 25 (50%) samples with Dengue fever, 4 (8%) samples with urinary tract infections, 3 (6%) samples with pneumonia, and 4 (8%) samples with local infection, 3 (6%) samples with sepsis and 11 (22%) samples with hepatitis B.

Comparison of microscopy, two RDTs and RT-PCR

The diagnostic value of Right Sign and ScreenPlus against microscopy consisted Sn = 100% and Sp = 98%, PPV = 98, 2%, NPV = 100% (Table 2). The results of the sensitivity and specificity of both tests on gold standard microscopy in this study obtained high, namely 100% and 98%. The results showed a good compatibility between RightSign, ScreenPlus and microscopy. The diagnostic value of RightSign and ScreenPlus against Real-Time PCR Sn = 75.3 % and Sp = 100 %, PPV = 100%, NPV = 64% (Table 2).

McNemar test showed that no significance different between RightSign with microscopic examination with P=1. There was a significant difference between RightSign and RT-PCR Results (Tables 3 and 4). Table 5 shows the variety of Plasmodium species detected by RightSign vs. microscopic and Table 6 showed the distribution of Plasmodium species against RT-PCR. RT-PCR detected more mix infection compared to microscopic, 24.8% vs. 2.9 %.

Tables 7 and 8 shows the comparison of the species of Plasmodium detected by ScreenPlus, microscopic and RT-PCR. The percentage was the same as the results of RightSign. McNemar statistical analysis on the performance differences between RightSign and ScreenPlus showed that there was no significant difference with P>0.05 (Tables 9 and 10). Mann Whitney analysis revealed no significant difference on parasitemia index between ScreenPlus and RightSign (P>0.05).

Discussion

The location for the sampling population in this study was Merauke, one of the

Table 1. Basic characteristics of subjects and study samples.

Variable	(n)	Total %
Total malaria	105	52.4
Positive microscopy test	54/105	51.42
Negative microscopy test	51/105	48.58
Positive rightsign test	54/105	51.42
Positive screenplus test	54/105	51.42
Parasitemia index <i>P. vivax</i> per μ L blood*	34/54	63
<1000	2/34	5.9
1000-10,000	18/34	52.9
10,001-200,000	14/34	41.2
>200,000	0/34	0
Parasitemia index <i>P. falciparum</i> per μ L blood*	17/54	31.5
<1000	1/17	5.9
1000-10,000	7/17	41.2
10,001-200,000	7/17	41.2
>200,000	2/17	11.7
Parasitemia index mixed (<i>P. falciparum</i> & <i>P. vivax</i>) per μ L blood*	3/54	5.5
<1000	-	-
1000-10,000	1/3	33.3
10,001-200,000	1/3	33.3
>200,000	1/3	33.3
Age (Mean \pm SD) years	0/3	0
Gender	31.19 \pm 17.97	-
Male	67/105	63.8
Female	38/105	36.2
Day of fever (Median (IQR)) days (n=28)	5 (2-60)	-

Table 2 Diagnostic performance of RightSign and ScreenPlus compared to microscopic and real-time PCR Examination.

Diagnostic performance	RightSign		ScreenPlus	
	Microscopic	RT-PCR	Microscopic	RT-PCR
Sn (%) (CI 95%)	100	75.3	100	75.3
Sp (%) (CI 95%)	98	100	98	100
PPV (%) (CI 95%)	98.2	100	98.2	100
NPV (%) (CI 95%)	100	64	100	64

Table 3. Comparison of RightSign vs. microscopic examination.

Right Sign	Microscopic results		Total	P-value*
	Positive	Negative		
Positive	54 (98.2%)	1 (1.8%)	55 (100%)	1.000
Negative	0 (0%)	50 (100%)	50 (100%)	
Total	54 (51.4%)	51	(48.6%)	105 (100%)

*Statistical analysis using McNemar test. N =105, significance at P<0.05.

Table 4. Comparison of RightSign vs. PCR results.

Right Sign	PCR Results		Total	P-value*
	Positive	Negative		
Positive	55 (100%)	0 (0%)	55 (100%)	< 0,001
Negative	18 (36%)	32 (64%)	50 (100%)	
Total	73 (69.5%)	32 (30.5%)	105 (100%)	

*Statistical analysis using McNemar test. N =105, significance at P<0.05.

districts in the Papua Province. Riskesdas in 2013 stated that Papua Province was one of the provinces with the highest malaria burden in Indonesia with Annual Parasites Incidence (API) of 45.85% in 2016. High malaria burden areas will affect the background of malaria exposure in the population, thus providing higher parasitic densities compared to non-endemic regional populations.⁴ Parasitic density will produce a high antigenemia protein resulting in high positivity in the detection of RDT protein antigens, as well as the positivity of Plasmodium detection by microscopy.

RDT sensitivity was influenced by various factors which included location and population sampling, antigenemia protein in the sample itself and optimal temperature stability of the reagent kit.^{13,14}

Optimal temperature stability influences Rapid diagnostic test (RDT) sensitivity (4-30°C). If the optimal temperature is stable, during delivery transportation or during examination, the RDT performance will run well and provide valid results.¹⁴ In this study, the temperature stability of the reagent kit was maintained at 4-30°C with the storage of reagent kits carried out in the refrigerator before being used. RDT sensitivity was also influenced by the suitability between observers in reading RDT results. Two observer concordance was analyzed by the McNemar test with P-value > 0.05 and no significant difference was found in the results. Arum I *et al.* showed similar results of RDT with this study. The diagnostic value of RDT against microscopy with Sn 100%, Sp 96.99%, PPV 83.2%, and NPV 100%.¹⁵

One false positive result of *P. falciparum* in RightSign and ScreenPlus in this study could be attributed to protein antigenemia HRP-2 malaria parasite that could still be identified in the blood of the patient for up to 30 days after antimalarial therapy due to slow clearance of HRP-2 in the blood. Another reason is the presence of gametocytes in the blood following antimalarial therapy. Gametocytes in the blood continue to produce all three antigenemia proteins (HRP-2, p-LDH and aldolase). Antigenemia p-LDH and aldolase proteins have faster clearance time from the blood after antimalarial therapy.^{16,17} Rheumatoid factor and heterophilic antibodies in the patient's blood are other cause of false positivity in the RDT.¹⁸

The results of the comparison between RDTs and gold standard microscopy examination were not statistically significantly different. The examiners involved in the microscopy examination may have influenced the results, however, the trained and certified malaria microscopy examiner

Table 5. Comparison of species detection between RightSign and microscopic.

Right Sign	Microscopic				Total
	Negative	<i>P. falciparum</i>	<i>P. vivax</i>	Mix	
Negative	50 (100%)	0 (0%)	0 (0%)	0 (0%)	50 (100%)
<i>P. falciparum</i>	1 (5.6%)	16 (88.9%)	0 (0%)	1 (5.6%)	18 (100%)
Pan	0 (0%)	1 (2.7%)	34 (91.9%)	2 (5.4%)	37 (100%)
Total	51 (48.6%)	17 (16.2%)	34 (32.4%)	3 (2.9%)	105 (100%)

Table 6. Comparison of species detection between RightSign and RT-PCR.

Right Sign	RT-PCR				Total
	Negative	<i>P. falciparum</i>	<i>P. vivax</i>	Mix	
Negative	32 (64%)	3 (6%)	10 (20%)	5 (10%)	50 (100%)
<i>P. falciparum</i>	0 (0%)	5 (27.8%)	0 (0%)	13 (72.2%)	18 (100%)
Pan	0 (0%)	0 (0%)	29 (78.4%)	8 (21.6%)	37 (100%)
Total	32 (30.5%)	8 (7.6%)	39 (37.1%)	26 (24.8%)	105 (100%)

Table 7. Comparison of species detection between ScreenPlus and microscopic results.

ScreenPlus	Microscopic examination				Total
	Negative	<i>P. falciparum</i>	<i>P. vivax</i>	Mix	
Negative	50 (100%)	0 (0%)	0 (0%)	0 (0%)	50 (100%)
<i>P. falciparum</i>	1 (5.9%)	17 (94.4%)	0 (0%)	0 (0%)	18 (100%)
<i>P. vivax</i>	0 (0%)	1 (2.7%)	34 (91.9%)	2 (5.4%)	37 (100%)
Total	51 (48.6%)	18 (16.2%)	34 (32.4%)	3 (2.9%)	105 (100%)

Table 8. Comparison of species detection between ScreenPlus and RT-PCR.

ScreenPlus	PCR Results				Total
	Negative	<i>P. falciparum</i>	<i>P. vivax</i>	Mix	
Negative	32 (64%)	3 (6%)	10 (20%)	5 (10%)	50 (100%)
<i>P. falciparum</i>	0 (0%)	5 (29.4%)	0 (0%)	13 (72.2%)	18 (100%)
<i>P. vivax</i>	0 (0%)	0 (0%)	29 (78.4%)	8 (21.6%)	37 (100%)
Total	32 (30.5%)	8 (7.6%)	39 (37.1%)	26 (24.8%)	105 (100%)

Table 9. Comparison of ScreenPlus and RightSign.

RightSign results	Screen plus results		Total	P-value*
	Positive	Negative		
Positive	55 (100%)	0 (0%)	55 (100%)	1,000
Negative	0 (0%)	50 (100%)	50 (100%)	
Total	55 (52.4%)	50 (47.6%)	105 (100%)	

*Statistical analysis using McNemar test. N = 105, significance at P < 0.05.

Table 10. Comparison of species detection between ScreenPlus and RightSign

RightSign Results	ScreenPlus Results			Total
	Negative	<i>P. falciparum</i>	<i>P. vivax</i>	
Negative	50 (100%)	0 (0%)	0 (0%)	50 (100%)
<i>P. falciparum</i>	0 (0%)	18 (94.4%)	0 (0%)	18 (100%)
Pan	0 (0%)	0 (0%)	37 (100%)	37 (100%)
Total	50 (47.6%)	17 (16.2%)	37 (35.2%)	105 (100%)

should produce a true and reliable results.^{10,12} Conformity between the results of RDTs and microscopy was also inseparable from the density of parasitemia. The population of malaria species in this study was dominated by *P. vivax* and then followed by *P. falciparum*, so that although there was a high parasitemia, but because of the dominance of malaria species occupied by *P. vivax*, the chance to produce false positives in pLDH was small. Patients with high *P. falciparum* parasitemia could give false positive results in pLDH and end with high *P. vivax* findings on RDT examination.¹⁹ The sensitivity of RightSign and ScreenPlus to RT-PCR was lower than the sensitivity to microscopy as the gold standard for malaria research, but the specificity is higher. A research in Flores reported that real-time PCR found that RT PCR was able to detect almost 8 times more cases of *Plasmodium* infection compared to microscopic examination as the gold standard for malaria testing. This is due to the ability of RT PCR to detect submicroscopic infections, with or without clinical malaria, that are not detected microscopically.^{10,20} This condition may be due to the ability of RT PCR to detect DNA fragments of malaria parasites when microscopically detecting the existence of *Plasmodium* malaria.²¹

Molecular methods are universally accepted to be more sensitive than microscopy, however PCR (multiplex PCR, real time PCR, or conventional) requires more sophisticated laboratories, trained personnel, longer times, and higher costs because PCR examination is not to be a routine examination of malaria. Malaria examination results can be more accurate if PCR is used as a reference standard for malaria testing because PCR method is able to detect parasitemia below the microscopic limit.

Conclusions

RightSign and ScreenPlus RDT have a very good performance. The analysis of *Plasmodium* antigen detection by RightSign, ScreenPlus, and microscopy research was not significantly different. Further research is needed to find out the diagnostic value of non-*falciparum* and non-*vivax Plasmodium* in RightSign and ScreenPlus. Real time-PCR detected higher mix infection when compared to microscopy and RDTs, therefore real-time PCR method can be considered as an effective confirmation tool for malaria diagnosis.

References

1. Surjadjaja C, Surya A, Baird JK. Epidemiology of *Plasmodium vivax* in Indonesia. *Am J Trop Med Hyg.* 2016; 95:121–32.
2. WHO. Indonesia South-East Asia Region - Malaria Profile. 2018; Available from: https://www.who.int/malaria/publications/country-profiles/profile_idn_en.pdf?ua=1
3. Tangpukdee N, Duangdee C, Wilairatana P, Krudsood S. Malaria diagnosis: A brief review. 2009;47:93–102.
4. Global WHO, Programme M. World Health Organization. Good practices for selecting and procuring rapid diagnostic tests for malaria. 2011: pp.1–71
5. Joanny F, Löhr SJ, Engleitner T, et al. Limit of blank and limit of detection of *Plasmodium falciparum* thick blood smear microscopy in a routine setting in Central Africa *Malar J* 2014;13:1–7.
6. Balaghaleh MR, Zarean M, Aghaee MA, et al. Comparison of PfHRP-2 / pLDH RDTs with Light Microscopy in a Low Prevalence Setting in Southeastern Iran, Sistan and Baluchestan: Due to Implementation of Malaria Elimination Program. 2018; 5:1–7.
7. Moody, A. Rapid diagnostic tests for malaria parasites. *Clin Microbiol Rev* 2002;15:66–78.
8. Berzosa P, De Lucio A, Romay-Barja M, et al. Comparison of three diagnostic methods (microscopy, RDT, and PCR) for the detection of malaria parasites in representative samples from Equatorial Guinea 11 Medical and Health Sciences 1108 Medical Microbiology. *Malar J* 2018;17:1–12. Available from: <https://doi.org/10.1186/s12936-018-2481-4>
9. WHO. Evaluations P. Recommended selection criteria for procurement of malaria rapid diagnostic tests. 2018.
10. Han TZ, Han KT, Aye KH, et al. Comparison of microscopy and PCR for the detection of human *Plasmodium* species and *Plasmodium knowlesi* in southern Myanmar. *Asian Pac J Trop Biomed* 2017;7:680–5. Available from: <http://dx.doi.org/10.1016/j.apjtb.2017.06.004>
11. Nakatsu M, Kato Y, Komaki-yasuda K, Perpe J. A novel PCR-based system for the detection of four species of human malaria parasites and *Plasmodium knowlesi* 2018;1–15.
12. World Health Organisation. Malaria microscopy quality assurance manual – Ver. 2. Who. 2016;140.
13. Boyle MJ, Wilson DW, Beeson JG. New approaches to studying *Plasmodium falciparum* merozoite invasion and insights into invasion biology. *Int J Parasitol* 2013;43:1–10.
14. Ashley EA, Touabi M, Ahrer M, et al. Evaluation of three parasite lactate dehydrogenase-based rapid diagnostic tests for the diagnosis of *falciparum* and *vivax* malaria. *Malar J* 2009;8:1–12.
15. Arum IL, Purwanto AP, Arfi S, et al. Uji diagnostik *Plasmodium* malaria menggunakan metode imunokromatografi diperbandingkan dengan pemeriksaan mikroskopis. *Indones J Clin Pathol Med Lab* 2018;12:118.
16. Hawkes M, Katsuva JP, Masumbuko CK. Use and limitations of malaria rapid diagnostic testing by community health workers in war-torn Democratic Republic of Congo. *Malar J* 2009;8:1–8.
17. Mogeni P, Williams TN, Omedo I et al. Detecting malaria hotspots: A comparison of rapid diagnostic test, microscopy, and polymerase chain reaction. *J Infect Dis.* 2017;216:1091–8.
18. Wongsrichanalai C, Barcus MJ, Muth S, et al. A review of malaria diagnostic tools: Microscopy and rapid diagnostic test (RDT). *Am J Trop Med Hyg* 2007;77:119–27.
19. Harris I, Sharrock WW, Bain LM, et al. A large proportion of asymptomatic *Plasmodium* infections with low and sub-microscopic parasite densities in the low transmission setting of Temotu Province, Solomon Islands: Challenges for malaria diagnostics in an elimination setting. *Malar J* 2010;9:1–8.
20. Khaimar K, Martin D, Lau R, et al. Multiplex real-time quantitative PCR, microscopy and rapid diagnostic immuno-chromatographic tests for the detection of *Plasmodium* spp: Performance, limit of detection analysis and quality assurance. *Malar J* 2009; 8:1–17.
21. Fransisca L, Kusnanto JH, Satoto TBT, et al. Comparison of rapid diagnostic test Plasmodium Malaria-3, microscopy, and quantitative real-time PCR for diagnoses of *Plasmodium falciparum* and *Plasmodium vivax* infections in Mimika Regency, Papua, Indonesia. 2015;1–11.

Performance comparison of two malaria rapid diagnostic test with real time polymerase chain reaction and gold standard of microscopy detection method

ORIGINALITY REPORT

16%

SIMILARITY INDEX

11%

INTERNET SOURCES

14%

PUBLICATIONS

0%

STUDENT PAPERS

PRIMARY SOURCES

- | | | |
|---|--|----|
| 1 | intj.infection.com
Internet Source | 1% |
| 2 | Roy Bagus Kurniawan, Puspa Wardhani, Heny Arwati, Aryati Aryati et al. "Association of Parasite Density and Hematological Parameters of Plasmodium vivax- and Plasmodium falciparum-infected Patients Attending Merauke General Hospital, Papua, Indonesia", Open Access Macedonian Journal of Medical Sciences, 2020
Publication | 1% |
| 3 | pure-oai.bham.ac.uk
Internet Source | 1% |
| 4 | Michael J. Kavanaugh, Steven E. Azzam, David M. Rockabrand. "Malaria Rapid Diagnostic Tests: Literary Review and Recommendation for a Quality Assurance, Quality Control Algorithm", Diagnostics, 2021
Publication | 1% |

5	<p>Wahyu Setyarini, Dadik Raharjo, Radita Yuniar Arizandy, Zakaria Pamoengkas et al. "Molecular detection of a new pathotype enteroaggregative haemorrhagic <i>Escherichia coli</i> (EAHEC) in Indonesia, 2015", <i>Infectious Disease Reports</i>, 2020</p> <p>Publication</p>	1 %
6	<p>pacmisc.net</p> <p>Internet Source</p>	1 %
7	<p>www.researchsquare.com</p> <p>Internet Source</p>	1 %
8	<p>malariajournal.biomedcentral.com</p> <p>Internet Source</p>	1 %
9	<p>link.springer.com</p> <p>Internet Source</p>	1 %
10	<p>www.alliedacademies.org</p> <p>Internet Source</p>	1 %
11	<p>qmro.qmul.ac.uk</p> <p>Internet Source</p>	<1 %
12	<p>Natanael Ritung, Victor D. Pijoh, Janno B. B. Bernadus. "Perbandingan Efektifitas Rapid Diagnostic Test (RDT) dengan Pemeriksaan Mikroskop pada Penderita Malaria Klinis di Puskesmas Mubune Kecamatan Likupang Barat", <i>Jurnal e-Biomedik</i>, 2018</p> <p>Publication</p>	<1 %

13 Xi Chen, Jiaqi Zhang, Maohua Pan, Yucheng Qin et al. "Loop-mediated isothermal amplification (LAMP) assays targeting 18S ribosomal RNA genes for identifying *P. vivax* and *P. ovale* species and mitochondrial DNA for detecting the genus *Plasmodium*", *Parasites & Vectors*, 2021
Publication

14 bmsd.com.my
Internet Source

15 repository.unair.ac.id
Internet Source

16 www.dovepress.com
Internet Source

17 www.nature.com
Internet Source

18 "Abstracts of the 18th Congress of the International Pediatric Nephrology Association, Venice (Italy), October 2019", *Pediatric Nephrology*, 2019
Publication

19 Phandee Watanaboonyongcharoen, Yara A. Park, Jessica L. Poisson, Mark E. Brecher. "Rapid increases in parasitemia following red cell exchange for malaria", *Journal of Clinical Apheresis*, 2011
Publication

20

Submitted to Universitas Airlangga

Student Paper

<1 %

21

Xiaodong, Sun, Ernest Tambo, Wei Chun, Cheng Zhibin, Deng Yan, Wang Jian, Wang Jiazhi, and Zhou Xiaonong. "Diagnostic performance of CareStart™ malaria HRP2/pLDH (Pf/pan) combo test versus standard microscopy on falciparum and vivax malaria between China-Myanmar endemic borders", Malaria Journal, 2013.

Publication

<1 %

22

Mahlet Belachew, Mistire Wolde, Desalegn Nega, Bokretsion Gidey et al. "Evaluating Performance of Multiplex Real Time PCR for the Diagnosis of Malaria at Elimination Targeted Low Transmission Settings of Ethiopia.", Research Square Platform LLC, 2021

Publication

<1 %

23

docplayer.net

Internet Source

<1 %

24

www.biopanda.co.uk

Internet Source

<1 %

25

Devi Oktafiani, Ni Luh Ayu Megasari, Elsa Fitriana, Nasronudin, Maria Inge Lusida, Soetjipto. "Human herpes virus 8 antibodies

<1 %

in HIV-positive patients in Surabaya,
Indonesia", *Infectious Disease Reports*, 2020

Publication

26

Ni Njoman Juliasih, Soedarsono, Reny Mareta Sari. "Analysis of tuberculosis program management in primary health care", *Infectious Disease Reports*, 2020

Publication

<1 %

27

wellcomeopenresearch.org

Internet Source

<1 %

28

worldwidescience.org

Internet Source

<1 %

29

Ariadna Martín-Díaz, José Miguel Rubio, Juan María Herrero-Martínez, Manolo Lizasoain et al. "Study of the diagnostic accuracy of microbiological techniques in the diagnosis of malaria in the immigrant population in Madrid", *Malaria Journal*, 2018

Publication

<1 %

30

Surasak Kasetsirikul, Jirayut Buranapong, Werayut Srituravanich, Morakot Kaewthamasorn, Alongkorn Pimpin. "The development of malaria diagnostic techniques: a review of the approaches with focus on dielectrophoretic and magnetophoretic methods", *Malaria Journal*, 2016

Publication

<1 %

- | | | |
|----|--|------|
| 31 | www.tandfonline.com
Internet Source | <1 % |
| 32 | www.update-software.com
Internet Source | <1 % |
| 33 | Mehrdad Ameri. "Laboratory diagnosis of malaria in nonhuman primates", <i>Veterinary Clinical Pathology</i> , 03/2010
Publication | <1 % |
| 34 | oro.open.ac.uk
Internet Source | <1 % |
| 35 | researchonline.lshtm.ac.uk
Internet Source | <1 % |
| 36 | Pragyan Dahal, Basudha Khanal, Keshav Rai, Vivek Kattel, Satish Yadav, Narayan Raj Bhattarai. "Challenges in laboratory diagnosis of Malaria in a low resource country among cases of acute febrile illness at tertiary care hospital in eastern Nepal: Comparative study on Conventional Vs Molecular approach", <i>Research Square</i> , 2020
Publication | <1 % |
| 37 | Regina Kandie, Rachel Ochola, Kariuki Njaanake. "Evaluation of fluorescent in-situ hybridization technique for diagnosis of malaria in Ahero Sub-County hospital, Kenya", <i>BMC Infectious Diseases</i> , 2018
Publication | <1 % |

Exclude quotes On

Exclude matches Off

Exclude bibliography On

Performance comparison of two malaria rapid diagnostic test with real time polymerase chain reaction and gold standard of microscopy detection method

GRADEMARK REPORT

FINAL GRADE

/100

GENERAL COMMENTS

Instructor

PAGE 1

PAGE 2

PAGE 3

PAGE 4

PAGE 5
