Validity of automated audiometry for hearing examination in patients with multidrug-resistant tuberculosis

by Nyilo Purnami

Submission date: 06-Apr-2022 05:58AM (UTC+0800)

Submission ID: 1802796150

File name: xamination in patients with multidrug-resistant tuberculosis.pdf (197.21K)

Word count: 4120 Character count: 22710



RESEARCH ARTICLE

Validity of automated audiometry for hearing examination in patients with multidrug-resistant tuberculosis [version 1; peer review: 1 approved]

Nyilo Purnami ¹⁰, Rian W. Palandeng¹, Soedarsono -², Dhany Arifianto³, In Seok Moon⁴

V1 First published: 14 Dec 2021, 10:1277 https://doi.org/10.12688/f1000research.75090.1 Latest published: 14 Dec 2021, 10:1277 https://doi.org/10.12688/f1000research.75090.1

Abstract 23

Background: The objective of this study was to test the validity of automated audiometry as a method of hearing examination in patients with multidrug-resistant tuberculosis.


Methods: This was a cross-sectional comparative study with a retrospective approach, using patient medical records. Patients with multidrug-resistant tuberculosis (MDR-TB) were recruited based-on medical records that met the inclusion and exclusion criteria at the Pulmonology outpatient unit, then referred to the Otorhinolaryngology outpatient unit of the Dr. Soetomo Academic Medical Center. The subjects' hearing function was measured with two different devices (automated audiometer and conventional audiometer) before being given anti-tuberculosis drug therapy (aminoglycoside injection) as ototoxicity monitoring from July to December 2019 period. Sensitivity and specificity analysis was used to assess the validity of the test.

Results: A total of 36 patients (72 ears) were included. The comparison test results using the Mann-Whitney test showed that there were significant differences between automated audiometry and conventional audiometry in both ears. Analysis values were: sensitivity 80-97%, specificity 37-96%, positive predictive value 74-98%, and negative predictive value 59-96%.

Conclusions: Automated audiometry is valid for use as a method of hearing examination and monitoring in patients with multidrugresistant tuberculosis.

Keywords

multidrug-resistant tuberculosis, ototoxicity, audiometry



¹Departement of Otorhinolaryngology Head and Neck Surgery, Airlangga University, Surabaya, Indonesia, -, Indonesia

²Department of Pulmonology and Respiratory Medicine, Airlangga University, Surabaya, Indonesia, -, Indonesia

³Departement of Engineering Physics, Institut Teknologi Sepuluh Nopember, Surabaya, Indonesia, -, Indonesia

⁴Department of Otorhinolaryngology, Yonsei University, Seoul, South Korea



Corresponding author: Nyilo Purnami (nyilo@fk.unair.ac.id)

Author roles: Purnami N: Conceptualization, Data Curation, Investigation, Project Administration; Palandeng RW: Resources,

Validation; - S: Data Curation, Resources; Arifianto D: Investigation, Validation; Moon IS: Resources

Competing interests: No competing interests were disclosed.

Grant information: This work was supported by Lembaga Pengelola Dana Pendidikan (LPDP) Rispro (Grant No.PRJ-80/LPDP/2019; to Soedarsono and Dhany Arifianto).

The funders had no role in study design, data collection and analysis, decision to publish, or preparation of the manuscript.

Copyright: © 2021 Purnami N *et al.* This is an open access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

How to cite this article: Purnami N, Palandeng RW, - S et al. Validity of automated audiometry for hearing examination in patients with multidrug-resistant tuberculosis [version 1; peer review: 1 approved] F1000Research 2021, 10:1277 https://doi.org/10.12688/f1000research.75090.1

First published: 14 Dec 2021, 10:1277 https://doi.org/10.12688/f1000research.75090.1

Introduction

Multidrug-resistant tuberculosis (MDR-TB) is tuberculosis (TB) resistant to isoniazid and rifampicin, with or without resistance to other anti-TB drugs. The World Health Organization (WHO) recommendations for multidrug-resistant tuberculosis (MDR-TB) include 8+ months of an aminoglycoside treatment such as kanamycin or amikacin or capreomycin. Aminoglycosides can produce significant side effects, including irreversible ototoxicity. The incidence of ototoxicity due to administration of aminoglycosides varies from 7% to 90%. Ototoxicity in MDR-TB patients is sensorineural and can be detected early by monitoring the hearing threshold periodically until the patient is recovered. Ototoxicity starts at high frequencies so that hearing techniques at high frequency are more sensitive to detect cochlear damage compared to methods that can only measure at standard frequencies (≤8000 Hz).

Initial hearing screening - at least air conduction (AC) - should be done on all patients who will start anti-tuberculosis drug therapy, especially aminoglycosides. Audiometry is a procedure to test one's listening ability at various sound frequencies and is used to identify hearing loss. This procedure is carried out using an electronic device called an audiometer to get the value of AC and bone conduction (BC). Not all audiometers can assess BC, so audiometry as auditory screening only requires AC values. MDR-TB patients with normal audiogram results can continue using anti-TB injections. ^{6,7}

Conventional audiometry is a gold standard examination to assess hearing loss. This procedure is carried out in a soundproof booth to determine the hearing threshold, which is the lowest pure tone that someone can still hear at a specific frequency, from 250 to 8000 Hz. The audiometer consists of a sound intensity control knob, a frequency control knob, headphones to assess AC and BC. Not all hospitals have soundproof chambers for this examination, and they are not recommended for MDR-TB patients because of the small size of the chamber. There is also less air circulation so they can cause shortness of breath and disturb concentration. 6.7

Automated audiometry is an audiometer device that, in its use, does not require a soundproof booth; or in other words, automated audiometry is a portable audiometer that can be used in an open space. There is an active noise monitoring feature that functions to monitor the high level of background noise when conducting audiometry, making it possible to pause the test until the background noise level returns to low. 9,10

Automated audiometry needs to be assessed for validity, and research in Indonesia has never been done. The purpose of this study was to prove the validity of automated audiometry as a method of hearing examination in patients with multidrug-resistant tuberculosis.

Methods

This research was a cross-sectional comparative study with a retrospective approach. The subjects of this study were patients with MDR-TB in the Pulmonology outpatient unit Dr. Soetomo Academic Medical Center, who were referred to the otorhinolaryngology outpatient clinic for examination of hearing function, before starting the anti-tuberculosis drug therapy (aminoglycoside injection) as monitoring of ototoxicity during the period from July to December 2019. Data were retrieved from medical records that met the inclusion and exclusion criteria. Inclusion criteria were new MDR-TB patients who performed two kinds of hearing examination using conventional audiometry as the gold standard of hearing assessment and automated audiometry which can measure at high frequencies. Exclusion criteria were patients with incomplete medical record data.

Automated audiometry uses the KUDUwave audiometer (model KUDUwave Prime), which can measure at frequencies from 250 Hz to 16 000 Hz. The patient uses headphones in an open space of the Pulmonology outpatient department with a noise level of 60 dB and is asked to press a button when a tone is heard. Conventional audiometry uses the Interacoustics AD226 audiometer, which can measure at frequencies of 125 Hz to 8000 Hz. The patient uses headphones in a soundproof chamber at the Pulmonology outpatient department with a noise level of 28 dB, and asked to press a button when a tone is heard. Calibration of the two audiometers is done routinely. Examination with automated audiometry and conventional audiometry from medical record data in this study was carried out by competent medical personnel.

Data obtained from the medical records included air conduction (AC) results from conventional audiometry and automated audiometry examinations. Other data recorded included sex, age, results of an otoscopy examination, pure tone average (PTA), and the degree of hearing loss based on ear count. The automated audiometry examination results were compared with conventional audiometry results that were calculated at all frequencies. The subsequent analysis with IBM SPSS Statistics 25.0 uses a 2×2 table, with the output in sensitivity, specificity, positive predictive value (PPV), and negative predictive value (NPV). Comparative analysis of automated audiometry and conventional audiometry using the Mann Whitney test.

Ethical clearance was obtained from the Ethical Committee of Dr. Soetomo Hospital, Surabaya, Indonesia (approval number 1858/KEPK/111.2020). Written informed consent was obtained from all individuals included in this study, after being given an explanation of the examinations to be carried out.

Results

Basic data

Based on data from medical records, the results of hearing tests using two methods were compared: (i) automated audiometry examination conducted in the open field of a Pulmonology outpatient department, (ii) conventional audiometry performed in a soundproof room as the gold standard of hearing function examination. Data were obtained from 36 patients (72 ears) in the study period. There were 21 male patients (58.33%) and 15 female patients (41.67%).

The youngest MDR-TB patient was 18 years old, while the oldest was 85 years old. The largest age group was 45 to 54 years, with 13 patients (36.11%). The results of the otoscopy examination in 36 patients (72 ears) showed all normal tympanic membranes (Table 1).

Conventional audiometry examination obtained normal hearing with an average of 19.26 ± 4.42 dB, mild hearing loss with an average of 29.52 ± 3.39 dB, moderate with an average of 45.62 ± 3.92 dB, moderate to severe with 62.50 ± 4.68 dB, and severe hearing loss with an average of 81.25 ± 12.37 dB (Table 2).

Automated audiometry examination results obtained normal hearing with an average of 16.93 ± 5.34 dB, mild hearing loss with an average of 31.67 ± 4.21 dB, moderate with an average of 50.78 ± 4.11 dB, moderate to severe degree with the average was 59.37 ± 0.88 dB and severe degree with an average of 87.50 ± 2.89 dB (Table 3).

The normality test results showed that the data were not normally distributed, so to find out significant differences between the two examinations, the Mann-Whitney test was used. The results showed significant differences (p < 0.05) between automated audiometry and conventional audiometry in both ears (Table 4).

Table 1. Distribution of the respondents.

13		
Age groups (years)	(n)	(%)
<15	0	0
15-24	4	11.11
25-34	4	11.11
35-44	10	27.78
45-54	13	36.11
55-64	4	11.11
≥65	1	2.78
<15	0	0
15-24	4	11.11
Total	36	100.00

Table 2. Conventional audiometry. SD = standard deviation.

Degree of hearing loss (pure tone average [PTA])	Mean	Median	SD
Normal (≤25 dB)	19.26	20.00	4.42
Mild (26-40 dB)	29.52	28.75	3.39
Moderate (41-55 dB)	45.62	45.00	3.92
Moderate to severe (56-70 dB)	62.50	62.50	4.68
Severe (71-90 dB)	81.25	81.25	12.37
Profound (≥91 dB)	-		-

Table 3. Automated audiometry. SD = standard deviation.

Degree of hearing loss (pure tone average [PTA])	Mean	Median	SD
Normal (≤ <mark>25 dB)</mark>	16.93	18.75	5.34
Mild (26-40 dB)	31.67	31.87	4.21
Moderate (41-55 dB)	50.78	51.87	4.11
Moderate to severe (56-70 dB)	59.37	59.37	0.88
Severe (71-90 dB)	87.50	87.50	2.89
Profound (≥91 dB)	-	-	-

Table 4. Comparative automated audiometry with conventional audiometry.

27	Pure tone average
Mann-Whitney U	2091.00
Wilcoxon W	4179.00
Z	-2.00
Sig. (2-tailed)	0.04

Table 5. Validity of automated audiometry. PPV = positive predictive value; NPV = negative predictive value.

	Degree of he	Degree of hearing loss				
	Normal	Mild	Moderate	Moderate to severe	Severe	Profound
Sensitivity	80%	89%	89%	97%	93%	-
Specificity	89%	37%	70%	85%	96%	-
PPV	95%	74%	86%	91%	98%	-
NPV	59%	64%	77%	96%	89%	-

The automated audiometry test results compared with conventional audiometry results as the gold standard, obtained a sensitivity of 80-97%, specificity 37-96%, positive predictive value (PPV) 74-98%, and negative predictive value (NPV) 59-96% (Table 5).

Discussion

The limitation of this study is that high frequencies (8000-16000 Hz) data collection of the automated audiometry was not carried out. The distribution of sex in this study found more male than female patients, consisting of 21 males (58.33%) and 15 females (41.67%). These results are consistent with research in China where 1154 MDR-TB incidents comprised 777 males and 377 females. MDR-TB is more frequent in males, a fact that is supported by research in Rawalpindi, Pakistan, that reports MDR-TB is more dominant in males with 23 cases than in females with 15 cases. However, a study in Ethiopia stated that the risk of MDR-TB decreases by 14% in males compared to females.

Another study in Surakarta reported MDR-TB cases in 50 males and 26 females. ¹⁴ The reason for this is not yet known, but could be due to male mobility or exposure due to social interactions is higher than female and non-compliance of a male patient in consuming anti-TB drugs. ¹⁵ A study about the risk of multidrug- or rifampicin-resistance in males versus females stated that there was no evidence of either sex being more at risk of MDR-TB. ¹⁶ The age characteristic of the youngest MDR-TB patients is 18 years, while the oldest is 85 years. The most populous age group was 45 to 54 years with 13 patients (36.11%). The average age of patients with MDR-TB was 43.44 years. Research in China reports that the most populous age group of MDR-TB patients is 31-45 years, with as many as 383 patients. ¹¹ Other studies in Mali report as many as 134 of 214 MDR-TB patients, including in the age group ≤40 years. ¹⁷ A study in Gujarat reported that majority of MDR-TB patients were aged between 40 to 50 years. ¹⁸ Age groups between 24-50 were found more in this study, probably because of its higher activity than other age groups.

The comparison test results using the Mann-Whitney test showed significant differences between automated audiometry and conventional audiometry in both ears. Research on the accuracy and efficiency of automated audiometry reports that automated audiometry is a stable, accurate, and time-efficient method for evaluating adult hearing status with normal hearing and hearing loss. 19 Research in South Africa stated that there is no significant difference between conventional audiometry and automated audiometry. 20 Several reports included in a systematic review indicated that automated audiometry using the method of adjustment (Békésy sweep or Békésy fixed frequency method) generally yields lower (i.e., better) thresholds compared with manual audiometry.2

Other studies report that conventional audiometry and audiometry hearing threshold results show a small difference.²⁷ Studies in primary school children aged 6-10 years in South Africa report that automated audiometry can correctly identify 87.5% of hearing loss detected using conventional audiometry.²⁸ Another study in industry reported that the difference in the hearing threshold between automated audiometry and conventional audiometry was less than 5 dB.2 The difference in the results of the two examinations in this study was probably due to the difference in the frequency of the two devices used and the different conditions (fatigue, shortness of breath) of patients with MDR-TB when examined.

The automated audiometry results against the conventional audiometry results obtained 80-97% sensitivity, specificity 37-96%, positive predictive value 74-98%, and negative predictive value 59-96%. Research evaluating the sensitivity and specificity of automated audiometry reports that automated audiometry has a high sensitivity, ranging from 86-100%and specificity of 78-100%. Positive predictive value is around 89-91%, and negative predictive value is about 89-100%, indicating that automated audiometry can be used to identify hearing loss.30 The results in this study were obtained according to the reference. The background noise level of a non-soundproofed room does not affect the accuracy of the hearing threshold value obtained using automated audiometry.

Conclusions

This study shows that automated audiometry is a valid method of hearing examination and monitoring in patients with multidrug-resistant tuberculosis with normal hearing or hearing loss. Automated audiometry does not require a soundproof booth, rather can be performed in an open space. An active noise monitoring feature monitors the high level of background noise when conducting audiometry, making it possible to pause the test until the background noise level returns to low.

Data availability

Underlying data

Figshare: Validity of automated audiometry for hearing examination in patients with multidrug-resistant tuberculosis. https://doi.org/10.6084/m9.figshare.17129123.3

Data are available under the terms of the Creative Commons Attribution 4.0 International license (CC-BY 4.0).

- Paul R: The threat of multidrug-resistant tuberculosis. J. Global Infect. Dis. 2018: 10(3): 119-120. PubMed Abstract | Publisher Full Text
- Heysell SK, Ahmed S, Rahman MT, et al.: Hearing loss with kanamycin treatment for multidrug-resistant tuberculosis in desh. Eur. Respir. J. 2018; 51(3): 1-3.
- Gamboa I, Sousa PC, Duarte D, et al.: Amikacin ototoxicity: case report and literature review. Otolaryngology Online Journal.
- Peterson L, Rogers C: Aminoglycoside-induced hearing deficits a review of cochlear ototoxicity. S. Afr. Fam. Pract. 2015; 57(2):
 - **Publisher Full Text**
- Ganesan P, Schmiedge J, Manchaiah V, et al.: Ototoxicity: a challenge in diagnosis and treatment. J. Audiol. Otol. 2018; 22(2): PubMed Abstract | Publisher Full Text
- Challenge TB: Guide to detect and manage hearing loss during the management of drug-resistant tuberculosis. 2019. Accessed February 1, 2020.

- Challenge TB: Audiometry in the management of drug-resistant tuberculosis, 2017, Accessed February 1, 2020, Reference Source
- American Speech Language Hearing Association: Guidelines for manual pure-tone threshold audiometry, 2019. Accessed Reference Source
- Shojaeemend H, Avatollahi H: Automated audiometry: a review of the implementation and evaluation methods. Healthc. Inform. Res. 2018; 24(4): 263–275. PubMed Abstract | Publisher Full Text
- Storey KK, Munoz K, Nelson L, et al.: Ambient noise impa accuracy of automated hearing assessment. Int. J. Audiol. 2014; 53(10): 730-736. PubMed Abstract | Publisher Full Text
- Lu Z, Jiang W, Zhang J, et al.: Drug resistance and epidemiology characteristics of multidrug-resistant tuberculosis patients in 17 provinces of China. PLoS One. 2019; 14(11): 1-14. Publisher Full Text
- Rafigue A. Dastgir M. Jamalullah M. et al.: Streptomycin associated hearing loss in patients treated for multidrug resistant tuberculosis. Isr. Med. J. 2012; 4(3): 139–142.

- Alemu A, Bitew ZW, Diriba G, et al.: Risk factors associated with drug-resistant tuberculosis in Ethiopia: a systematic review and meta-analysis. Authorea. 2021; 1: 1–12. Publisher Full Text
- Pamungkas P, Rahardjo SS, Murti B: Evaluation of multidrug resistant tuberculosis predictor index in Surakarta, Central Java. J. Epidemiol. Public Health. 2018; 03(2): 263–276. Publisher Full Text
- Soeroto AY, Pratiwi C, Santoso P, et al.: Factors affecting outcome of longer regimen multidrug-resistant tuberculosis treatment in West Java Indonesia: a retrospective cohort study. PLoS One. 2021; 16(2): 1–13.
 Publisher Full Text
- McQuaid CF, Horton KC, Dean AS, et al.: The risk of multidrug- or rifampicin-resistance in males versus females with tuberculosis. Eur. Respir. J. 2020; 56(3): 1–14.
- Baya B, Achenbach CJ, Kone B, et al.: Clinical risk factors associated with multidrug-resistant tuberculosis (MDR-TB) in Mali. Int. J. Infect. Dis. 2019; 81: 149–155.
 PubMed Abstract | Publisher Full Text
- Bhardwaj M, Shah KV: A cross-sectional study to correlate demographic and clinical profile of multi-drug resistant tuberculosis patients in a tertiary hospital. IP Indian J. Immunol. Respir. Med. 2021; 6(2): 80-85.
 Publisher Full Text
- Swanepoel DW, Mngemane S, Molemong S, et al.: Hearing assessment-reliability, accuracy, and efficiency of automated audiometry. Telemed. J. E Health. 2010; 16(5): 557-563. PubMed Abstract | Publisher Full Text
- Mahomed F, Swanepoel DW, Eikelboom RH, et al.: Validity of automated threshold audiometry: a systematic review and meta-analysis. Ear Hear. 2013; 34(6): 745–752.
 PubMed Abstract | Publisher Full Text
- Maiya PS, Kacker SK: Comparison of threshold between pure tone and Bekesy audiometry. Silent Wld. 1973; 8: 16–20.
- 22. Robinson DW, Whittle LS: A comparison of self-recording and manual audiometry: some systematic effects shown by

- unpractised subjects. J. Sound Vib. 1973; 26: 41–62. Publisher Full Text
- Erlandsson B, Häkanson H, Ivarsson A: Comparison of the hearing threshold measured by manual pue-tone and by self-recording (Békésy) audiometry. Audiology. 1979; 18: 414–429.
 PubMed Abstract | Publisher Full Text
- Harris DA: Microprocessor, self-recording and manual audiometry. J. Aud. Res. 1979; 19: 159–166.
 PubMed Abstract
- Frampton MC, Counter RT: A comparison of self recording audiometry in naval establishments and clinical audiometry in a hospital setting. J. Roy. Nav. Med. Serv. 1989; 75: 99-104. PubMed Abstract | Publisher Full Text
- IshakWS, Zhao F, Stephens D: Test-retest reliability and validity of Audioscan and Békésy compared with pure tone audiometry. Audiol. Med. 2011; 9: 40–46. Publisher Full Text
- Jones CGB, Eikelboom RH, Swanepoel DW, et al.: Clinical validation
 of automated audiometry with continuous noise-monitoring in
 a clinically heterogeneous population outside a soundtreated
 environment. Int. J. Audiol. 2016; 55(9): 507–513.
 PubMed Abstract | Publisher Full Text
- Mahomed AF, Swanepoel DW, Eikelboom RH: Diagnostic hearing assessment in schools: validity and time efficiency of automated audiometry. J. Am. Acad. Audiol. 2016; 27: 42–48.
- Meinke DK, Norris JA, Flynn BP, et al.: Going wireless and boothless for hearing testing in industry. Int. J. Audiol. 2017; 56: 41–51. PubMed Abstract | Publisher Full Text
- Skjonsberg A, Heggen C, Jamil M, et al.: Sensitivity and specificity
 of automated audiometry in subjects with normal hearing or
 hearing impairment. Noise & Health A Bimonthly Inter-disciplinary
 International Journal. 2019; 21(98): 1–6.
 Publisher Full Text
- Purnami N: Validity of automated audiometry for hearing examination in patients with multidrug-resistant tuberculosis. figshare. Dataset. 2021.
 Publisher Full Text

Open Peer Review

Current Peer Review Status:





Reviewer Report 02 February 2022

https://doi.org/10.5256/f1000research.78907.r115235

© 2022 Jenkins H. This is an open access peer review report distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.



Herman A. Jenkins

Department of Otolaryngology, University of Colorado School of Medicine, Aurora, CO, USA

Access and therefore expense of providing standard audiometry testing is often difficult, particularly in developing countries with areas of limited access. This manuscript compares standard audiometry with an automated form. The latter does not require sound dampened facilities or great expense. The investigators compared the two and demonstrated that, while the automated testing was less predictive of hearing level, it was within a range to make it feasible for situations not allowing sophisticated setup.

The authors have used a small, but valid sample and the conclusions that they have drawn are valid. Monitoring in conditions as drug resistance or chemotherapy is important and the use of automated testing provided sufficient information to validate the hearing level without requiring expensive set up. I would recommend indexing the manuscript as is.

Is the work clearly and accurately presented and does it cite the current literature? Yes

Is the study design appropriate and is the work technically sound?

Vas

Are sufficient details of methods and analysis provided to allow replication by others?

If applicable, is the statistical analysis and its interpretation appropriate?

Are all the source data underlying the results available to ensure full reproducibility?

Yes

Are the conclusions drawn adequately supported by the results?

Yes

Competing Interests: No competing interests were disclosed.

Reviewer Expertise: I am Chair of an Otolaryngology-Head and Neck Surgery Department in the USA with expertise in hearing and balance disorders and monitoring both.

I confirm that I have read this submission and believe that I have an appropriate level of expertise to confirm that it is of an acceptable scientific standard.

Author Response 03 Feb 2022

Nyilo Purnami, Airlangga University, Surabaya, Indonesia

Thank you very much.

Competing Interests: No competing interest

The benefits of publishing with F1000Research:

- · Your article is published within days, with no editorial bias
- · You can publish traditional articles, null/negative results, case reports, data notes and more
- The peer review process is transparent and collaborative
- · Your article is indexed in PubMed after passing peer review
- · Dedicated customer support at every stage

For pre-submission enquiries, contact research@f1000.com



Validity of automated audiometry for hearing examination in patients with multidrug-resistant tuberculosis

ORIGINA	ALITY REPORT			
SIMILA	9 RITY INDEX	15% INTERNET SOURCES	15% PUBLICATIONS	0% STUDENT PAPERS
PRIMARY	Y SOURCES			
1	reposito	ory.zums.ac.ir		1 %
2	journals Internet Sour	.plos.org		1 %
3	www.ch	allengetb.org		1 %
4	reposito	ory.up.ac.za		1 %
5	global.o	•		1 %
6	WWW.NC	bi.nlm.nih.gov		1 %
7		Commentary: D more problems 009		0/2
8	journals Internet Sour	.sagepub.com		1 %

scifes.fkm.ui.ac.id

1 %

Jessica van Tonder, De Wet Swanepoel, Faheema Mahomed-Asmail, Hermanus Myburgh, Robert H. Eikelboom. "Automated Smartphone Threshold Audiometry: Validity and Time Efficiency", Journal of the American Academy of Audiology, 2017

<1%

Sarah Y. Bessen, Isabelle L. Magro, Karen Mojica Alvarez, Devin R. Cowan et al. "Test-Retest repeatability of automated threshold audiometry in Nicaraguan schoolchildren", International Journal of Audiology, 2022

<1%

www.eurosurveillance.org

<1%

Benjamin D. Thumamo Pokam, D. Yeboah-Manu, P.M. Teyim, P.W. Guemdjom, B. Wabo, A.B.D. Fankep, R.E. Okonu, Anne E. Asuquo. "A pilot study on the genetic diversity of Mycobacterium tuberculosis complex strains from tuberculosis patients in the Littoral region of Cameroon", Journal of Clinical

Tuberculosis and Other Mycobacterial Diseases, 2020

Publication

15	purehost.bath.ac.uk Internet Source	<1%
16	WWW.cureus.com Internet Source	<1%
17	Faheema Mahomed, De Wet Swanepoel, Robert H. Eikelboom, Maggi Soer. "Validity of Automated Threshold Audiometry", Ear & Hearing, 2013	<1%
18	cyberleninka.org Internet Source	<1%
19	repozitorij.mef.unizg.hr Internet Source	<1%
20	Ayinalem Alemu, Zebenay Workneh Bitew, Getu Diriba, Balako Gumi. "Risk factors associated with drug - resistant tuberculosis in Ethiopia: A systematic review and meta - analysis", Transboundary and Emerging Diseases, 2021	<1%
21	Hilman Pardede, Kalamullah Ramli, Yohan Suryanto, Nur Hayati, Alfan Presekal. "Speech	<1%

Enhancement for Secure Communication

Using Coupled Spectral Subtraction and Wiener Filter", Electronics, 2019 Publication

- researchonline.lshtm.ac.uk <1% 22 Internet Source journalotohns.biomedcentral.com 23 Internet Source Ping Wang, Bo Wang, Xinhui Chen, Bing Xiong, < 1 % 24 Fangyao Xie, Sheng Wu, Yinshan Tang, Shuqi Chen, Xueping Ding, Peng Liu, Wei Luo. "Six-Year Follow-Up of Dysphagia in Patients with Parkinson's Disease", Dysphagia, 2021 Publication Resom Berhe Gebremariam, Maereg Wolde, <1% 25 Aykerem Beyene. "Determinants of adherence to anti-TB treatment and associated factors among adult TB patients in Gondar city administration, Northwest, Ethiopia: based on health belief model perspective", Journal of Health, Population and Nutrition, 2021 Publication fisika.fst.unair.ac.id <1% Internet Source
- ir.unilag.edu.ng
 Internet Source

A. Chaudhary, T. Mahmood, A.D. Shukla, A. Shreenivasa, Verma Arvind, K. Ahmad, A. Verma. "Association of Socio-Demographic Profile with Prevalence of Multi Drug Resistant Tuberculosis among Retreated Pulmonary Tuberculosis Patients in North India", SAARC Journal of Tuberculosis, Lung Diseases and HIV/AIDS, 2018

<1%

Publication

Boye Bobby Carter, Yang Zhang, Hangjin Zou, Chuhan Zhang, Xinming Zhang, Rongtian Sheng, Yanfei Qi, Changgui Kou, Yin Li. "Survival analysis of patients with tuberculosis and risk factors for multidrug-resistant tuberculosis in Monrovia, Liberia", PLOS ONE, 2021

<1%

Publication

Fei Zhao, Dafydd Stephens, Mark Rogers, Peter Harper. "Audiological investigations in patients with late onset facioscapulohumeral muscular dystrophy (FSHD)", Audiological Medicine, 2009

<1%

Publication

Soedarsono Soedarsono, Tutik Kusmiati, Ariani Permatasari. "Effect of Diabetes Mellitus on Renal and Audiology Toxicities in

Patients with Drug-Resistant Pulmonary Tuberculosis", Archives of Clinical Infectious Diseases, 2021

Publication

33	Tiantian Zhang, Wei Liu, Li Li, Zou Jue, chunhua xu. "Evaluation of serum and pleural levels tumor M2-pyruvate kinase in lung cancer patients with pleural effusion", Research Square Platform LLC, 2022 Publication	<1%
34	etd.repository.ugm.ac.id Internet Source	<1%
35	ipindexing.com Internet Source	<1%
36	learningfundamentals.com Internet Source	<1%
37	De Wet Swanepoel. "Hearing Assessment— Reliability, Accuracy, and Efficiency of Automated Audiometry", Telemedicine Journal and e-Health, 06/2010 Publication	<1%
38	M. Dody Izhar, Marta Butar Butar, Fajrina Hidayati, Ruwayda Ruwayda. "Predictors and health-related quality of life with short form- 36 for multidrug-resistant tuberculosis patients in Jambi, Indonesia: A case-control	<1%

study", Clinical Epidemiology and Global Health, 2021

Publication

Meenakshi R. Ramanathan, Crystal K. Howell, James M. Sanders. "Drugs in tuberculosis and leprosy", Elsevier BV, 2019

Publication

<1%

- i ubiicatioi
- Qinglin Cheng, Li Xie, Le Wang, Min Lu, Qingchun Li, Yifei Wu, Yinyan Huang, Qingjun Jia, Gang Zhao. "Incidence Density and Predictors of Multidrug-Resistant Tuberculosis Among Individuals With Previous Tuberculosis History: A 15-Year Retrospective Cohort Study", Frontiers in Public Health, 2021

<1%

Alessandra G Samelli, Camila M Rabelo, Seisse G Gandolfi Sanches, Ana C Martinho, Carla G Matas. "Tablet-based tele-audiometry:
Automated hearing screening for schoolchildren", Journal of Telemedicine and

<1%

Publication

Telecare, 2018

42

Brennan-Jones, Christopher G., Robert H. Eikelboom, De Wet Swanepoel, Peter L. Friedland, and Marcus D. Atlas. "Clinical validation of automated audiometry with continuous noise-monitoring in a clinically heterogeneous population outside a sound-

treated environment", International Journal of Audiology, 2016.

Publication

43

Marize Brittz, Barbara Heinze, Faheema Mahomed-Asmail, De Wet Swanepoel, Anton Stoltz. "Monitoring Hearing in an Infectious Disease Clinic with mHealth Technologies", Journal of the American Academy of Audiology, 2020

<1%

Publication



Seong-Su Lee, Kyung-do Han, Young-Hoon Joo. "Association of perceived tinnitus with duration of hormone replacement therapy in Korean postmenopausal women: a crosssectional study", BMJ Open, 2017

<1%

Publication

Exclude quotes

On

Exclude matches

Off

Exclude bibliography On

Validity of automated audiometry for hearing examination in patients with multidrug-resistant tuberculosis

GRADEMARK REPORT	
final grade /100	GENERAL COMMENTS Instructor
PAGE 1	
PAGE 2	
PAGE 3	
PAGE 4	
PAGE 5	
PAGE 6	
PAGE 7	
PAGE 8	
PAGE 9	