

# Imprinted Zeolite Modified Carbon Paste Electrode as a Selective Sensor for Blood Glucose Analysis by Potentiometry

*by Miratul Khasanah*

---

**Submission date:** 15-Dec-2020 05:04PM (UTC+0800)

**Submission ID:** 1475642959

**File name:** Glucose\_IZ-LTA\_IJC\_2020.pdf (410.92K)

**Word count:** 5281

**Character count:** 26654

## Imprinted Zeolite Modified Carbon Paste Electrode as a Selective Sensor for Blood Glucose Analysis by Potentiometry

Miratul Khasanah\*, Alfa Akustia Widati, Usreg Sri Handajani, Muji Harsini, Bahrotul Ilmiah, and Irene Dinda Oktavia

Department of Chemistry, Faculty of Science and Technology, Universitas Airlangga, Campus C, Jl. Dr. Ir. H. Soekarno (MERR), Surabaya 60115, Indonesia

### \* Corresponding author:

tel: +62-8123077233  
email: miratul-k@fst.unair.ac.id

Received: September 19, 2019  
Accepted: January 30, 2020

DOI: 10.22146/ijc.49820

**Abstract:** Imprinted zeolite modified carbon paste (carbon paste-IZ) electrode had been developed as a sensor to analyze blood glucose content by potentiometry. The used zeolite was Lynde Type A (LTA) that synthesized with a mole ratio of  $\text{Na}_2\text{O}$ ,  $\text{Al}_2\text{O}_3$ ,  $\text{SiO}_2$  and  $\text{H}_2\text{O}$  of 4:1:1.8:270, respectively while non-imprinted zeolite was prepared with a mole ratio of glucose/Si of 0.0306. Glucose was then extracted from the zeolite framework using hot water (80 °C) to produce imprinted zeolite (IZ). The carbon paste-IZ electrode prepared from activated carbon, paraffin pastilles, and IZ with a mass ratio of 5:4:1 showed the best performance. The modified electrode demonstrated the measurement range of  $10^{-4}$ - $10^{-2}$  M, the Nernst factor of 29.55 mV/decade, the response time less than 120 s, and the detection limit of  $5.62 \times 10^{-5}$  M. Ascorbic acid, uric acid, urea and creatinine did not interfere on the glucose analysis by potentiometry. Comparison test with spectrophotometry showed an accuracy of  $(90.7 \pm 1.4)\%$  ( $n = 5$ ), while the application of the electrode to analyze five spiked serum samples showed recovery of  $(92.2 \pm 1.3)\%$  ( $n = 5$ ). The electrode was stable for up to 9 weeks (168 times usage). Based on its performance, the developed electrode can be applied to analyze glucose in human serum sample and recommended for used in the medical field.

**Keywords:** blood glucose; carbon paste electrode; imprinted zeolite; potentiometric sensor

## ■ INTRODUCTION

Diabetes mellitus has become a serious threat to human health because it can cause kidney failure, heart disease, and stroke. An increase in high blood glucose levels ( $> 200$  mg/dL) is a major cause of diabetes mellitus [1]. In the body, normal blood glucose values are 60–100 mg/dL ( $3.3 \times 10^{-3}$ – $5.5 \times 10^{-3}$  M), while serum glucose is 70–110 mg/dL ( $3.88 \times 10^{-3}$ – $6.1 \times 10^{-3}$  M) [2]. Based on the effects that can be caused by high blood glucose levels, controlling the glucose level in the blood is necessary. The general method for determining glucose levels is spectrophotometry using chemical or enzymatic, such as hexokinase or glucose 6-phosphate dehydrogenase [3]. A few studies have reported about the use of some methods for glucose determination including GOx (glucose oxidase) based colorimetric methods [4-5], high performance liquid chromatography (HPLC) [6], liquid chromatography-

mass spectrometry (LC-MS) [7], and potentiometry using modified carbon-based electrodes [8]. Modified carbon-based electrodes have previously been developed including zeolite-modified electrodes for the analysis of uric acid [9],  $\text{Cr}^{3+}$  ions [10], creatinine [11], creatine [12]; and polymer-based electrodes for phenol analysis [13]. Nanoparticle material has also been shown to increase the selectivity and sensitivity of electrodes to analytes [14].

In this study, an imprinted zeolite modified carbon paste electrode has been developed for potentiometric glucose detection. The zeolite used was LTA zeolite that synthesized from the basic materials of  $\text{SiO}_2$ ,  $\text{NaAlO}_2$ , and water with a mole ratio of  $\text{Na}_2\text{O}$ ,  $\text{Al}_2\text{O}_3$ ,  $\text{SiO}_2$ ,  $\text{H}_2\text{O}$  of 4:1:1.8:270 [15]. Glucose solution was added to the mixture of these materials to produce non imprinted zeolites (NIZ). Next, the glucose molecule was extracted, so that it leaves a mold on the zeolite framework called

imprinted zeolite (IZ). The imprinted is expected only match the size and shape to glucose molecules so that they can selectively detect glucose molecules when it is used for sample analysis.

Electrode performance was studied from the measurement range value, Nernst factor, detection limit, response time and the life time of the electrode. The use of electrodes was studied through the accuracy value compared with the spectrophotometry method as a standard method for determining blood glucose levels in the medical field. Electrode selectivity was investigated through the effect of another component in serum samples including ascorbic acid, uric acid, creatinine, and urea on the glucose analysis using the electrode.

## EXPERIMENTAL SECTION

### Materials

Chemicals used in this study were glucose (Sigma Aldrich, 99.5%), SiO<sub>2</sub> (Sigma Aldrich, Ludox 40%), sodium aluminate (Sigma Aldrich, 50%), sodium dihydrogen phosphate (Merck, 99%), disodium hydrogen phosphate (Merck, 99.5%), uric acid (Fluka, 98%), urea, creatinine, H<sub>3</sub>PO<sub>4</sub>, AgNO<sub>3</sub>, Ag wire, *n*-hexane, paraffin pastilles and activated carbon with pore size of 3.835 nm. The sample used was blood serum from a clinical laboratory in Surabaya.

### Instrumentation

The equipment used in this study was a Cyberscan 510 potentiometers with Ag/AgCl as a reference electrode, Fourier transform infrared (FTIR) spectrophotometer (Shimadzu IR Prestige 21), X-ray diffraction (PANalytical Expert Pro), Spectrophotometer (UV 1800 Pharmaspec), hotplate (Termolyne S46410-2), pH meter (Cyberscan Eutech pH 510), centrifuge (HITTECH EBA) 20, vacuum oven (NAPCO Model 5851), magnetic stirrer, polypropylene bottle, agate mortar, 1000 µL micropipette tip and glassware.

### Procedure

#### Synthesis of zeolite, non-imprinted zeolite (NIZ), and imprinted zeolite (IZ)

The LTA zeolite was synthesized by mixing Na<sub>2</sub>O, Al<sub>2</sub>O<sub>3</sub>, SiO<sub>2</sub>, and H<sub>2</sub>O with a mole ratio of 4:1:1.8:270 [15].

The mixture was stirred for 3 h. A one third part was heated in an oven at 100 °C for 45 h. Next, it is washed using distilled water with the help of a centrifuge. The remaining 2/3 parts of the mixture were added with glucose to produce a glucose/Si mole ratio of 0.0306. The obtained mixture was allowed to stand for 3 h so that glucose can be binded and trapped in the zeolite framework. A half part of the mixture was dried at 80 °C and the other part was added with hot water to extract glucose and form imprinted zeolite.

#### Fabrication of carbon paste-imprinted zeolite and pH optimization

The carbon paste electrodes were prepared by inserting Ag wires in a 1 mL micropipette tube. Three quarter part of the micropipette tip was filled with melted paraffin. Then a paste made from a mixture of activated carbon, solid paraffin, and IZ with various compositions (Table 1) was filled into the ¼ part remaining of the micropipette tube. The surface of the electrodes was rubbed on the HVS paper and immersed in a 10<sup>-2</sup> M glucose solution overnight for conditioning.

Furthermore, each electrode was used to measure the glucose standard solution 10<sup>-8</sup>–10<sup>-1</sup> M pH 6 (without pH adjustment). Then a graph between the concentration and electrode potential was created and determined the value of the Nernst factor and linear measurement range. The composition of carbon paste-IZ electrodes that produce optimum performance is used as the basis for preparing zeolite-modified carbon paste electrodes (EZ) and NIZ-modified carbon paste electrodes (ENIZ). The performance of EZ and ENIZ were compared with carbon paste-IZ to observe the effect of glucose mold on the electrode performance. The effect of pH solution was

**Table 1.** Composition of activated carbon, paraffin, and IZ on electrode fabrication

Electrode	% weight		
	Activated carbon	paraffin	IZ
E1	60	40	0
E2	55	40	5
E3	50	40	10
E4	45	40	15
E5	40	40	20

studied through the Nernst factor value and linear measurement range resulting from the measurement of glucose solution with varies pH.



#### Performance of the electrode and method validity

Parameters for expressing electrode performance and method validity in this study include linear measurement range, Nernst factor, detection limit, precision, accuracy, response time, life time and selectivity. The selectivity of the electrode was stated by the selectivity coefficient ( $K_{ij}$ ) value determined by the matched potential method (MPM) [16]. The accuracy was determined through a comparison test with the results of blood glucose analysis using the spectrophotometric method as a commonly used method in the medical field.

## RESULTS AND DISCUSSION

### Synthesis and Characterization of Zeolite, NIZ, and IZ

The LTA zeolite is prepared through a sol-gel method using silica sol as the silica source and sodium aluminate as the aluminium source. The precursor solution is combined under vigorous stirring. The mixture is then sealed in polypropylene bottles for the hydrothermal process. The reactions that occur in zeolite synthesis are as shown in Eq. (1) and (2).



$\text{SiO}_2$  was added to  $\text{NaAl}(\text{OH})_4$  solution dropwise and stirred for 1 h until it became a white gel. This process is an aging step which is the process of forming a zeolite crystal core [18]. The aging process aims the breaking of silica monomers into a stronger silica gel with the expected pore [19].

Furthermore, the mixture was heated hydrothermally at  $100^\circ\text{C}$  for 45 h in a polypropylene bottle to accelerate the growth of zeolite crystals [20]. From this process, a white suspension is produced. A total of 1/3 parts of the suspension are centrifuged at 4000 rpm for 5 min. Then the precipitate was separated and washed with distilled water to remove the remaining NaOH and dried it in an oven at  $80^\circ\text{C}$  for 24 h.

Non imprinted zeolite (NIZ) is zeolite which contains templates/mold in the pores. Non imprinted zeolite was prepared by taking as much as 2/3 parts of a mixture of  $\text{SiO}_2$ ,  $\text{NaAlO}_2$ , and  $\text{H}_2\text{O}$  (after the hydrothermal process). Then the mixture was added with glucose solution dropwise and stirred with a magnetic stirrer for




Fig 1. An illustration of trapping glucose molecules into a zeolite framework


30 min. An illustration of trapping glucose molecules into a zeolite framework can be seen in Fig. 1. The imprinted zeolite (IZ) was obtained by extracting glucose molecules from the NIZ framework using hot water. Benedict tests were performed to ensure that all glucose has been extracted from the NIZ framework (Fig. 2).

Characterization of zeolites by XRD was carried out to determine the formation of synthesized LTA zeolites. Fig. 3 is an X-ray diffraction pattern of LTA zeolites.

Based on the XRD pattern of Fig. 3, there is a peak with high intensity at position  $2\theta$  i.e. 7.23; 10.16; 12.46; 16.09; 21.67; 23.97; 27.11; 29.93; and 34.17°. At position  $2\theta$  12.46°, it shows the orientation of the cubic structure of the LTA zeolite [22]. Furthermore, these peaks were compared with the LTA zeolite diffraction pattern in the Collection of Simulated XRD Powder Patterns for Zeolites International Zeolite Association (IZA) data base [21] and American Society for Testing and Materials (ASTM) standard no. 39-222 [23]. There is a similarity between the peaks of the synthesized zeolite with the zeolite diffractogram in the data base. (Table 2) It can be concluded that synthesized LTA zeolite has formed. The peak at  $2\theta$  13.79° which is a typical peak of silicon dioxide ( $\text{SiO}_2$ ), indicates that there is remaining unreacted  $\text{SiO}_2$  to form zeolites.

Characterization with FTIR spectrophotometers produced the spectrum in Fig. 4. The wave number  $1640\text{ cm}^{-1}$  shows the peak of bending vibrations -OH


from Si-OH, Al-OH, and  $\text{H}_2\text{O}$  and the peak at wave number  $990\text{ cm}^{-1}$  is the asymmetry stretching vibrations of Al-O and Si-O [24]. The peak at wave number  $561\text{ cm}^{-1}$



**Fig 2.** Benedict test on (a) glucose standard solution, (b) leached filtrate of NIZ, a (c) leached filtrate of IZ

**Table 2.** The data peak position of synthesized LTA zeolite and data base from IZA and ASTM

2 $\theta$ (°)		
ASTM	Synthesized zeolite	IZA
7.18	7.23	7.18
10.16	10.16	10.17
12.45	12.46	12.46
-	13.79	-
16.09	16.09	16.11
21.65	21.67	21.67
23.97	23.97	23.99
27.09	27.11	27.11
29.92	29.93	29.94
34.16	34.17	34.18



**Fig 3.** XRD-pattern of synthesized LTA zeolite and LTA zeolite on IZA [21]



indicates the crystallization of zeolites with a double ring [25] and the peak at wave number  $440\text{ cm}^{-1}$  is a tetrahedral internal vibration of Si-O and Al-O [24]. The wavenumber at  $3460\text{ cm}^{-1}$  in non imprinted zeolite spectrum is broader than the peak at zeolite or imprinted zeolite. This is due to the presence of stretching -OH from C-OH in glucose [26].

### Optimization of Electrode Composition and pH Solution

The carbon paste-IZ electrodes are prepared from a mixture of activated carbon, paraffin and IZ. The presence of imprinted zeolite is expected to increase the electrode selectivity to glucose because IZ has a selective mold for glucose. The composition of activated carbon and IZ in the manufacture of electrodes may affect the performance of the electrodes, therefore the preparation of electrodes with variations in the composition of carbon and IZ were conducted. The resulted measurement range and Nernst factor from the measuring data of glucose solution using the carbon paste-IZ electrode with varies composition are shown in Table 3.

The electrode that exhibits a Nernstian graph on the electrode potential versus  $\log [\text{glucose}]$  plot is the E3 (Table 3). The E3 electrodes also show good linearity in the measurement range of  $10^{-4}$ - $10^{-2}$  M. The carbon paste modified zeolite (EZ) and NIZ (ENIZ) were prepared by composition of activated carbon, paraffin and zeolite or NIZ similar to the composition of the E3 electrode to determine the effect of glucose mold on the electrode performance. NIZ-modified electrodes produce curves close to Nernstian, while zeolite-modified electrodes

show the shape of sub-Nernstian curves. It is suspected that the presence of glucose molecules in NIZ enhances the electrodes to recognize free glucose molecules in solution. Thus, ENIZ functions like an ion selective electrode in recognizing glucose molecule.

The IZ modified electrode with the best performance was then used to measure the glucose solutions  $10^{-8}$ - $10^{-1}$  M at pH 4, 5, 6, 7, and 8. Measurement data was used to create a graph of  $\log [\text{glucose}]$  versus




Fig 4. FTIR spectra of glucose, zeolite LTA, NIZ, and IZ

Table 3. Value of the electrode measurement range, Nernst factor and linearity of calibration graph in the glucose measurement (without pH adjustment)

Electrode	Nernst factor (mV/decade)	Linear measurement range (M)	Linearity ( $R^2$ )
E1	27.79	$10^{-4}$ - $10^{-2}$	0.9549
E2	27.71	$10^{-4}$ - $10^{-2}$	0.9512
E3	29.55	$10^{-4}$ - $10^{-2}$	0.9924
E4	24.55	$10^{-4}$ - $10^{-2}$	0.9434
E5	16.16	$10^{-4}$ - $10^{-2}$	0.9604
EZ*	13.50	$10^{-4}$ - $10^{-2}$	0.9720
ENIZ*	27.35	$10^{-4}$ - $10^{-2}$	0.9147

\*) similar composition with E3

electrode potential (Fig. 5) and determined the measurement range, linearity and the Nernst factor (Table 4).

Fig. 5 explained that the best performance of E3 is shown in the measurement of glucose solution without adjusting pH (pH 6). From these measurements obtained linear and Nernstian curve with the equation of  $y = 29.55x + 175.52$  in the concentration range of  $10^{-4}$ – $10^{-2}$  M and non-linear curve with the equation of  $y = -0.1x^2 - 3.18x + 38.43$ . Based on the intersection of both lines, the detection limit value obtained was  $5.62 \times 10^{-5}$  M. The value is lower than the detection limit of the conductive polymer-based sensors developed previously [8]. Table 5 displayed the comparison data of the limit of detection (LOD) of our work with the previously reports.


With this low detection limit, carbon paste-IZ electrodes can be used for glucose analysis with concentrations up to 100 times lower than normal concentrations in the blood [2]. The linear curve describes

the linear measurement range of the potentiometric method using the developed electrodes, which is  $10^{-4}$ – $10^{-2}$  M with a Nernst factor of 29.55 mV/decade. It shows that glucose has two valency molecules. This result is in agreement with the research report of [31] which fabricated the potentiometric glucose sensor based on glucose oxidase immobilized iron ferrite magnetic particle/chitosan composite modified gold coated glass electrode. They resulted in the Nernst factor of glucose about 27.3 mV/decade. A similar result also reported by [32] which stated glucose is a divalent molecule in the analysis of glucose using the voltammetry method. The valency of the molecule is constant in both analyses using potentiometry and voltammetry. The comparison of potentiometry and voltammetry method in the  $\text{Ca}^{2+}$  sensor is used as an analogue approach, whereas both of the methods found a similar Nernst factor value [33-34].

**Table 4.** Data of electrode measurement range, Nernst factor, and linearity on varies pH of glucose solution

pH	Measurement range (M)	Nernst factor (mV/decade)	Linearity ( $R^2$ )
4	$10^{-3}$ – $10^{-1}$	1.40	0.8547
5	$10^{-7}$ – $10^{-4}$	2.49	0.9208
6	$10^{-6}$ – $10^{-4}$	1.25	0.9868
6*	$10^{-4}$ – $10^{-2}$	29.55	0.9924
7	$10^{-4}$ – $10^{-1}$	3.02	0.8341
8	$10^{-6}$ – $10^{-4}$	3.35	0.9964

\*) pH of glucose solution without pH adjustment



**Fig 5.** The plot of log [glucose] and electrode potential on varies pH

**Table 5.** The comparison data of the LOD of our work with the previously reports

Method	The material of the electrode	Linear range	LOD	Reference
Amperometry	Carbon paste/GOx silica	$5 \times 10^{-4}$ – $9 \times 10^{-3}$ M	$1.5 \times 10^{-4}$ M	[27]
Amperometry	Carbon paste/selenium nanoparticle-mesoporous silica composite (MCM-41)	$1 \times 10^{-5}$ – $2 \times 10^{-3}$ M	$1 \times 10^{-4}$ M	[28]
Potentiometry	Poly (3-aminophenyl boronic acid-co-3-octylthiophene)	$5 \times 10^{-3}$ – $5 \times 10^{-2}$ M	$5 \times 10^{-4}$ M	[8]
Potentiometry	Poly (terthiophene benzoic acid) (pTBA) layered-AuZn alloy oxide (AuZnOx)	$1.6 \times 10^{-3}$ – $2.7 \times 10^{-2}$ M	$9.6 \times 10^{-4}$ M	[29]
Potentiometry	Carbon nanotube on gold printed	$10^{-3}$ – $10^{-1}$ M	$1 \times 10^{-4}$ M	[30]
Potentiometry	Carbon paste/IZ	$10^{-4}$ – $10^{-2}$ M	$5.6 \times 10^{-5}$ M	This work

The coefficient of variation (CV) in the linear range is 0.42–2.20%. This value is within the limit of the CV value required by the Association of Analytical Chemist (AOAC) which is 5.3% for concentrations less than  $10^{-2}$  M [35].

#### The Selectivity of the Electrode

Electrode selectivity to glucose molecules is expressed by the value of the selectivity coefficient ( $K_{ij}$ ). This value is calculated using the matched potential method (MPM) [16]. Determination of  $K_{ij}$  is begun by measuring separately the glucose, ascorbic acid, uric acid, creatinine and urea solution with a concentration range of  $10^{-4}$ – $10^{-2}$  M using bare carbon paste and carbon paste-IZ electrode, then the selectivity coefficient value of each electrode is determined. The  $K_{ij}$  values of bare carbon paste and carbon paste-IZ electrode are shown in Table 6.

Data on Table 6 illustrates that the presence of ascorbic acid, uric acid, creatinine or urea did not interfere with glucose potentiometric analysis using imprinted zeolite modified carbon paste electrodes, while the bare carbon paste electrodes are disturbed by these components. It explains that the modification of electrodes using imprinted zeolite can increase electrode selectivity, due to the presence of specific molds for glucose molecules in the zeolite. The selectivity of carbon paste-IZ electrode is as good as the zeolite-based electrodes developed previously [11-12,36].

#### Method Comparison Test and Recovery

The stability of electrode was studied using a comparison test of the results of the analysis with the potentiometric method using a carbon paste-IZ electrode with a spectrophotometric method which is

**Table 6.** Data of  $K_{ij}$  value of bare carbon paste electrode and carbon paste-IZ electrode

Solution	Concentration (M)	Selectivity coefficient ( $K_{ij}$ )	
		Bare carbon paste	Carbon paste-IZ
Ascorbic acid	$10^{-4}$	1.66	$3.24 \times 10^{-3}$
	$10^{-3}$	2.59	$3.48 \times 10^{-3}$
	$10^{-2}$	3.67	$3.79 \times 10^{-3}$
Uric acid	$10^{-4}$	1.98	$2.27 \times 10^{-3}$
	$10^{-3}$	2.70	$2.64 \times 10^{-3}$
	$10^{-2}$	4.98	$3.71 \times 10^{-3}$
Creatinine	$10^{-4}$	2.21	$1.62 \times 10^{-2}$
	$10^{-3}$	3.01	$1.77 \times 10^{-2}$
	$10^{-2}$	4.76	$1.94 \times 10^{-2}$
Urea	$10^{-4}$	2.21	$5.21 \times 10^{-3}$
	$10^{-3}$	3.19	$6.25 \times 10^{-3}$
	$10^{-2}$	4.68	$9.37 \times 10^{-3}$



**Table 7.** Data of accuracy and recovery value on the determination of blood serum glucose

Solution	Glucose concentration (M)		Accuracy (%)	Recovery (%)
	Potentiometry	Spectrophotometry*)		
Glucose $10^{-3}$ M	$8.38 \times 10^{-4}$	-	-	-
Serum sample 1	$4.19 \times 10^{-3}$	$4.55 \times 10^{-3}$	92.02	-
Serum sample 1 + Glucose $10^{-3}$ M	$4.97 \times 10^{-3}$	-	-	93.4
Serum sample 2	$4.18 \times 10^{-3}$	$4.72 \times 10^{-3}$	88.60	-
Serum sample 2 + Glucose $10^{-3}$ M	$4.94 \times 10^{-3}$	-	-	90.3
Serum sample 3	$5.85 \times 10^{-3}$	$6.38 \times 10^{-3}$	91.70	-
Serum sample 3 + Glucose $10^{-3}$ M	$6.66 \times 10^{-3}$	-	-	92.5
Serum sample 4	$8.11 \times 10^{-3}$	$8.89 \times 10^{-3}$	91.20	-
Serum sample 4 + Glucose $10^{-3}$ M	$8.89 \times 10^{-3}$	-	-	93.10
Serum sample 5	$11.09 \times 10^{-3}$	$12.22 \times 10^{-3}$	90.72	-
Serum sample 5 + Glucose $10^{-3}$ M	$11.86 \times 10^{-3}$	-	-	92.18

\*) data from clinical laboratory

standard method for determining blood glucose levels in the medical field. Sampling and handling of blood samples were carried out by a clinical laboratory in Surabaya. The recovery test is carried out to determine the influence of the matrix in the serum sample through standard addition techniques. Data on the accuracy and recovery resulted in blood glucose analysis are shown in Table 7.

Based on Table 7, the accuracy value obtained from the analysis of blood serum samples in this study is the range of accuracy received for a chemical analysis method according to AOAC, which is 80–110% [35]. This shows that the potentiometric method to analyze glucose in serum using the developed electrodes give in good agreement results from spectrophotometric method. Recovery value in this study illustrates that there are other components contained in serum samples that interfere with glucose analysis such as fructose, sucrose, maltose [37] and ions such as  $\text{Na}^+$ ,  $\text{Cl}^-$ ,  $\text{K}^+$ ,  $\text{PO}_4^{3-}$ ,  $\text{SO}_4^{2-}$ ,  $\text{NH}_4^+$  [38]. However, based on the obtained recovery value, the developed potentiometric method fulfill the requirements as a chemical analysis method. Thus, the developed method can be used as an alternative method for determining blood glucose levels in the medical field.

## CONCLUSION

The carbon paste-IZ electrodes for potentiometric glucose analysis showed a measurement range of  $10^{-4}$ – $10^{-2}$  M, Nernst factor 29.55 mV/decade, and able to detect concentrations up to 100 times lower than normal

concentrations of glucose in the blood serum. The electrode exhibited fast response and high selectivity against glucose molecules in solutions containing ascorbic acid, uric acid, creatinine, or urea. Potentiometry using developed electrodes shows high accuracy toward spectrophotometry as a standard method for blood glucose analysis in the medical field. Based on its performance, the potentiometric method using carbon paste-IZ electrodes is recommended as an alternative method for the analysis of blood glucose levels in the medical field.

## ACKNOWLEDGMENTS

The authors thank the Ministry of Research, Technology and Higher Education, Indonesia for the financial support of this study through PDUPT Grant No. 713/UN3.14/LT/2019 Universitas Airlangga and for the Chemistry Department Universitas Airlangga for the laboratory facilities provided.

## REFERENCES

- [1] Tonyushkina, K., and Nichols, J.H., 2009, Glucose meters: A review of technical challenges to obtaining accurate results, *J. Diabetes Sci. Technol.*, 4 (3), 971–980.
- [2] Kee, J.L., Hayes, E.R., and McCuiston, L.E., 2015, *Pharmacology: A Patient Centered Nursing Process Approach*, 8<sup>th</sup> Ed., Elsevier Saunders, St. Louis.
- [3] Bishop, D.K., La Belle, J.T., Vossler, S.R., Patel, D.R., and Cook, C.B., 2010, A disposable tear

- glucose biosensor-Part 1: Design and concept testing, *J. Diabetes Sci. Technol.*, 4 (2), 299-306.
- [4] Yoo, E.H., and Lee, S.Y., 2010, Glucose biosensors: An overview of use in clinical practice, *Sensors*, 10 (5), 4558-4576.
- [5] Galant, A.L., Kaufman, R.C., and Wilson, J.D., 2015, Glucose: Detection and analysis, *Food Chem.*, 188, 149-160.
- [6] Odden, J., 2011, Determination of D-glucose in human blood serum using HPLC-PED, *Concordia Coll. J. Anal. Chem.*, 2, 58-66.
- [7] Kamal, A.M., and Klein, P., 2011, Determination of sugar in honey by liquid chromatography, biomedical, and applied sciences, *Saudi J. Biol. Sci.*, 18 (1), 17-21.
- [8] Çiftçi, H., Tamer, U., Teker, M.Ş., and Pekmez, N.Ö., 2013, An enzyme free potentiometric detection of glucose based on a conducting polymer poly (3-aminophenyl boronic acid-co-3-octylthiophene), *Electrochim. Acta*, 90, 358-365.
- [9] Khasanah, M., Harsini, M., and Widati, A.A., 2013, Imprinting zeolite-modified glassy carbon as a voltammetric sensor for uric acid, *Indones. J. Chem.*, 13 (2), 108-113.
- [10] Heidari, Z., and Masrournia, M., 2018, A novel modified carbon paste electrode for the determination of chromium(III) in water, *J. Anal. Chem.*, 78 (8), 824-831.
- [11] Khasanah, M., Handajani, U.S., Widati, A.A., Abdulloh, A., and Rindarti, R.R., 2018, Construction and performance of creatinine selective electrode based on carbon paste-imprinting zeolite, *Anal. Bioanal. Electrochem.*, 10 (4), 429-438.
- [12] Athiroh, A., Fadillah, T., Damayanti, D.F., Abdulloh, A., Widati, A.A., and Khasanah, M., 2019, Carbon paste electrode modified imprinted zeolite as a selective sensor for creatine analysis by potentiometry, *IOP Conf. Ser.: Earth Environ. Sci.*, 217, 012003.
- [13] Rahmadhani, S., Setiyanto, H., and Zulfikar, M.A., 2018, Fabrication of carbon paste electrode modified with phenol imprinted polyaniline as a sensor for phenol analysis by potentiometric, *Mater. Sci. Forum*, 936, 71-76.
- [14] Jiang, L.C., and Zhang, W.D., 2010, A highly sensitive non-enzymatic glucose sensor based on CuO nanoparticles-modified carbon nanotube electrode, *Biosens. Bioelectron.*, 25 (6), 1402-1407.
- [15] Pera-Titus, M., Bausach, M., Lorens, J., and Cunnill, F., 2008, Preparation of inner-side tubular zeolite NaA membranes in a continuous flow system, *Sep. Purif. Technol.*, 59 (2), 141-150.
- [16] Tohda, K., Dragoe, D., Shibata, M., and Umezawa, Y., 2001, Studies on matched potential method for determining selectivity coefficients of ion-selective electrode based on neutral ionophores: Experimental and theoretical verification, *Anal. Sci.*, 17 (6), 733-743.
- [17] Baerlocher, C.H., McCusker, L.B., and Olson, D.H., 2007, *Atlas of Zeolite Framework Types*, 6<sup>th</sup> Ed., Elsevier Science, Amsterdam.
- [18] Houssin, C.J.Y., 2003, Nanoparticles in Zeolite Synthesis, *Dissertation*, Technische Universiteit Eindhoven, Netherlands.
- [19] Smitha, S., Shajesh, P., Aravind, P.R., Kumar, S.R., Pillai, P.K., and Warriar, K.G.K., 2006, Effect aging time and concentration of aging solution on the porosity characteristic of subcritically dried silica aerogels, *Microporous Mesoporous Mater.*, 91 (1-3), 286-292.
- [20] Cundy, C.S., and Cox, P.A., 2005, The hydrothermal synthesis of zeolites: Precursors, intermediates and reaction mechanism-review, *Microporous Mesoporous Mater.*, 82 (1-2), 1-78.
- [21] Treacy, M.M.J., and Higgins, J.B., 2001, *Collection of Simulated XRD Powder Patterns for Zeolites*, 4<sup>th</sup> Ed., Elsevier Science, Amsterdam.
- [22] Huang, A., Wang, N., and Caro, J., 2012, Synthesis of multi-layer zeolite LTA membranes with enhanced gas separation performance by using 3-aminopropyl triethoxysilane as interlayer, *Microporous Mesoporous Mater.*, 164, 294-301.
- [23] Selim, M.M., and Abd El-Maksoud, I.H., 2004, Hydrogenation of edible oil over zeolite prepared

- from local kaolin, *Microporous Mesoporous Mater.*, 74 (1-3), 79–85.
- [24] Rios, C.A., Williams, C.D., and Fulen, M.A., 2009, Nucleation and growth history of zeolite LTA synthesized from kaolinite by two different methods, *Appl. Clay Sci.*, 42 (3-4), 446–454.
- [25] Alkan, M., Hopa, C., Yilmaz, Z., and Guler, H., 2005, The effect of alkali concentration and solid/liquid ratio on the hydrothermal synthesis of zeolite NaA from natural kaolinite, *Microporous Mesoporous Mater.*, 86 (1-3), 176–184.
- [26] Silverstein, R.M., Webster, F.X., and Kiemle, D.J., 2005, *Spectrometric Identification of Organic Compounds*, 7<sup>th</sup> Ed., John Wiley & Sons, New York.
- [27] Jedrzak, A., Rębiś, T., Klapiszewski, L., Zdarta, J., Milczarek, G., and Jesionowski, T., 2018, Carbon paste electrode based on functional GOx/silica-lignin system to prepare an amperometric glucose biosensor, *Sens. Actuators, B*, 256, 176–185.
- [28] Yusan, S., Rahman, M.M., Mohamad, N., Arrif, T.M., Latif, A.Z.A., Mohd Aznan, M.A., and Wan Nik, W.S.B., 2018, Development of an amperometric glucose biosensor based on the immobilization of glucose oxidase on the Se-MCM-41 mesoporous composite, *J. Anal. Methods Chem.*, 2018, 2687341.
- [29] Kim, D.M., Cho, S.J., Cho, C.H., Kim, K.B., Kim, M.Y., and Shim, Y.B., 2016, Disposable all-solid-state pH and glucose sensors based on conductive polymer covered hierarchical AuZn oxide, *Biosens. Bioelectron.*, 79, 165–172.
- [30] Alhans, R.A., Singh, A., Singhal, C., Narang, J., Wadhwa, S., and Mathur, A., 2018, Comparative analysis of single-walled and multi-walled carbon nanotubes for electrochemical sensing of glucose on gold printed circuit boards, *Mater. Sci. Eng., C*, 90, 273–279
- [31] Khun, K., Ibupoto, Z.H., Lu, J., AlSalhi, M.S., Atif, M., Ansari, A.A., and Wilander, M., 2012, Potentiometric glucose sensor based on the glucose oxidase immobilized iron ferrite magnetic particle/chitosan composite modified gold coated glass electrode, *Sens. Actuators, B*, 173, 698–703.
- [32] Park, S., Boo, H., and Chung, T.D., 2006, Electrochemical non-enzymatic glucose sensors, *Anal. Chim. Acta*, 556 (1), 46–57.
- [33] Mousavi, M.P.S., Ainla, A., Tan, E.K.W., Abd El-Rahman, M.K., Yoshida, Y., Yuan, L., Sigursslid, H.H., Arkan, N., Yip, M.C., Abrahamson, C.K., Homer-Vanniasinkam, S., and Whitesides, G.M., 2018, Ion sensing with thread-based potentiometric electrodes, *Lab Chip*, 18 (15), 2279–2290.
- [34] Zhang, J., Harris, A.R., Cattral, R.W., and Bond, A.M., 2010, Voltammetric ion-selective electrodes for the selective determination of cations and anions, *Anal. Chem.*, 82 (5), 1624–1633.
- [35] Taverniers, I., De Loose, M., and Van Bockstaele, E., 2004, Trends in quality in the analytical laboratory. II. Analytical method validation and quality assurance, *TrAC, Trends Anal. Chem.*, 23 (8), 535–552.
- [36] Khasanah, M., Harsini, M., Widati, A.A., and Ibrani, P.M., 2017, The influence of ascorbic acid, creatine, and creatinine on the uric acid analysis by potentiometry using a carbon paste modified imprinting zeolite electrode, *J. Chem. Technol. Metall.*, 52 (6), 1039–1044.
- [37] Amani-Beni, Z., and Nezamzadeh-Ejhieh, A., 2017, A novel non-enzymatic glucose sensor based on the modification of carbon paste electrode with CuO nanoflower: Designing the experiments by response surface methodology (RSM), *J. Colloid Interface Sci.*, 504, 186–196.
- [38] Guyton, A.C., and Hall, J.E., 2006, *Textbook of Medical Physiology*, 11<sup>th</sup> Ed., Elsevier, Philadelphia.

# Imprinted Zeolite Modified Carbon Paste Electrode as a Selective Sensor for Blood Glucose Analysis by Potentiometry

## ORIGINALITY REPORT

18%

SIMILARITY INDEX

8%

INTERNET SOURCES

16%

PUBLICATIONS

0%

STUDENT PAPERS

## PRIMARY SOURCES

- 1** Miratul Khasanah, Handoko Darmokoesoemo, Nesti Widayanti, Yassine Kadmi, Hicham Elmsellem, Heri Septya Kusuma. **3%**  
"Development of carbon paste electrodes modified by molecularly imprinted polymer as potentiometry sensor of uric acid", Results in Physics, 2017  
Publication
- 2** dl.uctm.edu **2%**  
Internet Source
- 3** Miratul Khasanah, Handoko Darmokoesoemo, Nunung Mareta Sari, Yassine Kadmi, Hicham Elmsellem, Heri Septya Kusuma. **2%**  
"Development of electrode carbon paste modified by molecularly imprinted polymer as sensor for analysis of creatinine by potentiometric", Results in Physics, 2017  
Publication
- 4** Miratul Khasanah, Handoko Darmokoesoemo, Lendhy Kustyarini, Yassine Kadmi, Hicham **1%**

Elmsellem, Heri Septya Kusuma. "RETRACTED: Development of electrode carbon paste/molecularly imprinted polymer (MIP) with methacrylic acid as monomer to analyze glucose by potentiometry", Results in Physics, 2017

Publication

5

M. Khasanah, H. Darmokoesoemo, D. A. Rizki. "Carbon paste electrode modified molecularly imprinted polymer as a sensor for creatinine analysis by stripping voltammetry", AIP Publishing, 2017

Publication

1 %

6

Philippe Bühlmann, Ernö Pretsch, Eric Bakker. "Carrier-Based Ion-Selective Electrodes and Bulk Optodes. 2. Ionophores for Potentiometric and Optical Sensors", Chemical Reviews, 1998

Publication

1 %

7

[agris.fao.org](http://agris.fao.org)

Internet Source

<1 %

8

[www.degruyter.com](http://www.degruyter.com)

Internet Source

<1 %

9

Ferial Nosratinia, Hazhir Ghahremani, Saeed Shirazian. "Preparation and characterization of nanoporous ceramic membranes for separation of water from ethanol", Desalination and Water Treatment, 2014

<1 %



10

Zhu, D.. "Corrosion protection of AA 2024-T3 by bis-[3-(triethoxysilyl)propyl]tetrasulfide in sodium chloride solution.", Corrosion Science, 200310

Publication

<1 %

11

Aerogels Handbook, 2011.

Publication

<1 %

12

Jordan, Trace. "Chemistry", Oxford University Press

Publication

<1 %

13

Yang, C.. "Determination of tetrandrine and fangchinoline in plasma samples using hollow fiber liquid-phase microextraction combined with high-performance liquid chromatography", Journal of Chromatography A, 20070914

Publication

<1 %

14

Mongia Said Zina. "ESR study of MoY and PdMoY reduction", Magnetic Resonance in Chemistry, 03/2004

Publication

<1 %

15

[open.uct.ac.za](http://open.uct.ac.za)

Internet Source

<1 %

16

MADHURI LAKHANE, RAJENDRA KHAIRNAR, MEGHA MAHABOLE. "Metal oxide blended

<1 %

ZSM-5 nanocomposites as ethanol sensors",  
Bulletin of Materials Science, 2016

Publication

17

[aip.scitation.org](http://aip.scitation.org)

Internet Source

<1 %

18

Castells, C.B.. "Effect of temperature on pH measurements and acid-base equilibria in methanol-water mixtures", Journal of Chromatography A, 20030620

Publication

<1 %

19

Luo, Liqiang, Limei Zhu, and Zhenxin Wang. "Nonenzymatic amperometric determination of glucose by CuO nanocubesâ€“graphene nanocomposite modified electrode", Bioelectrochemistry, 2012.

Publication

<1 %

20

Marco Morigi, Erika Scavetta, Mario Berrettoni, Marco Giorgetti, Domenica Tonelli. "Sulfate-selective electrodes based on hydrotalcites", Analytica Chimica Acta, 2001

Publication

<1 %

21

[worldwidescience.org](http://worldwidescience.org)

Internet Source

<1 %

22

Ahmed E. Abdelhamid, Ahmed A. El-Sayed, Ahmed M. Khalil. "Polysulfone nanofiltration membranes enriched with functionalized graphene oxide for dye removal from

<1 %

23

Hilal Torul, Hakan Çiftçi, Demet Çetin, Zekiye Suludere, Ismail Hakkı Boyacı, Uğur Tamer. "Paper membrane-based SERS platform for the determination of glucose in blood samples", Analytical and Bioanalytical Chemistry, 2015

Publication

---

<1 %

24

Kamireddy Sirisha, Sreedhar Mallipattu, Sreenivasula Reddy Jayarama Reddy. "Differential Pulse Adsorptive Stripping Voltammetric Determination of Chlorpyrifos at a Sepiolite Modified Carbon Paste Electrode", Analytical Letters, 2007

Publication

---

<1 %

25

Rahman, Md. Mahbubur, A. J. Saleh Ahammad, Joon-Hyung Jin, Sang Jung Ahn, and Jae-Joon Lee. "A Comprehensive Review of Glucose Biosensors Based on Nanostructured Metal-Oxides", Sensors, 2010.

Publication

---

<1 %

26

Rothenberg. "Heterogeneous Catalysis", Catalysis, 01/01/2008

Publication

---

<1 %

27

Vlastimil Vyskocil, Jiri Barek. "Electroanalysis of Nitro and Amino Derivatives of Polycyclic

<1 %

# Aromatic Hydrocarbons", Current Organic Chemistry, 2011

Publication

---

28	<a href="http://bdigital.unal.edu.co">bdigital.unal.edu.co</a> Internet Source	<1 %
29	<a href="http://en.wikipedia.org">en.wikipedia.org</a> Internet Source	<1 %
30	<a href="http://epdf.pub">epdf.pub</a> Internet Source	<1 %
31	<a href="http://eprints.lincoln.ac.uk">eprints.lincoln.ac.uk</a> Internet Source	<1 %
32	<a href="http://jurnal.ugm.ac.id">jurnal.ugm.ac.id</a> Internet Source	<1 %
33	<a href="http://link.springer.com">link.springer.com</a> Internet Source	<1 %
34	<a href="http://sinta3.ristekdikti.go.id">sinta3.ristekdikti.go.id</a> Internet Source	<1 %
35	<a href="http://www.semanticscholar.org">www.semanticscholar.org</a> Internet Source	<1 %
36	Jiang, L.C.. "A highly sensitive nonenzymatic glucose sensor based on CuO nanoparticles-modified carbon nanotube electrode", Biosensors and Bioelectronics, 20100215 Publication	<1 %

---

37

Miratul Khasanah, Handoko Darmokoesoemo, Nunung Mareta Sari, Yassine Kadmi, Hicham Elmsellem, Heri Septya Kusuma. "RETRACTED: Development of electrode carbon paste modified by molecularly imprinted polymer as sensor for analysis of creatinine by potentiometric", Results in Physics, 2017

<1 %

Publication

---

Exclude quotes Off

Exclude matches Off

Exclude bibliography On



# Imprinted Zeolite Modified Carbon Paste Electrode as a Selective Sensor for Blood Glucose Analysis by Potentiometry

---

## GRADEMARK REPORT

---

FINAL GRADE

**/0**

GENERAL COMMENTS

**Instructor**

---

PAGE 1

---

PAGE 2

---

PAGE 3

---

PAGE 4

---

PAGE 5

---

PAGE 6

---

PAGE 7

---

PAGE 8

---

PAGE 9

---

PAGE 10

---